

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101525\
 Data File : VX048191.D
 Acq On : 15 Oct 2025 15:17
 Operator : JC/MD
 Sample : Q3348-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WCS-TP1

Quant Time: Oct 16 04:06:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 10/17/2025
 Supervised By : Mahesh Dadoda 10/17/2025

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	120180	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	214794	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	247536	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	129128	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	87903	45.951	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	91.900%		
35) Dibromofluoromethane	5.373	113	85871	59.611	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	119.220%		
50) Toluene-d8	8.635	98	252357	50.628	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	101.260%		
62) 4-Bromofluorobenzene	11.061	95	123663	65.371	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	130.740%#		
Target Compounds						
						Qvalue
16) Acetone	2.380	43	15190	18.262	ug/l	91
18) Methyl Acetate	2.703	43	6855	3.703	ug/l	100
20) Methylene Chloride	2.788	84	6401	3.636	ug/l	84
25) 2-Butanone	4.568	43	7115	6.858	ug/l	96
37) Ethyl Acetate	4.708	43	4643m	2.060	ug/l	
43) Isopropyl Acetate	6.330	43	7069	1.912	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101525\
Data File : VX048191.D
Acq On : 15 Oct 2025 15:17
Operator : JC/MD
Sample : Q3348-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WCS-TP1

Quant Time: Oct 16 04:06:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 10/17/2025
Supervised By :Mahesh Dadoda 10/17/2025

