

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101525\
 Data File : VX048193.D
 Acq On : 15 Oct 2025 15:58
 Operator : JC/MD
 Sample : Q3346-01 500X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Oct 16 04:07:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	96445	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	203702	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	217427	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	111435	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	96844	63.084	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	126.160%#
35) Dibromofluoromethane	5.373	113	76586	56.060	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	112.120%
50) Toluene-d8	8.635	98	266908	56.463	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	112.920%#
62) 4-Bromofluorobenzene	11.067	95	111096	61.926	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	123.860%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101525\
Data File : VX048193.D
Acq On : 15 Oct 2025 15:58
Operator : JC/MD
Sample : Q3346-01 500X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :

Quant Time: Oct 16 04:07:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

