

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101618\
 Data File : VX005358.D
 Acq On : 16 Oct 2018 09:40
 Operator : JC/MD
 Sample : J5521-15
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-2018-10-13-1

Quant Time: Oct 17 05:19:46 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	278048	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	384870	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	349583	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	171430	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	161460	52.32	ug/l	-0.01
Spiked Amount	50.000		Recovery	=	104.64%	
35) Dibromofluoromethane	5.50	113	125333	49.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.34%	
50) Toluene-d8	8.71	98	467041	53.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.74%	
62) 4-Bromofluorobenzene	11.14	95	152768	46.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.94%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101618\
Data File : VX005358.D
Acq On : 16 Oct 2018 09:40
Operator : JC/MD
Sample : J5521-15
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-2018-10-13-1

Quant Time: Oct 17 05:19:46 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration

