

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101618\
 Data File : VX005361.D
 Acq On : 16 Oct 2018 11:00
 Operator : JC/MD
 Sample : J5521-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TWP-01

Quant Time: Oct 17 05:36:40 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	275543	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	397938	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	363898	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	192283	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	155927	50.99	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.98%	
35) Dibromofluoromethane	5.50	113	124935	47.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.80%	
50) Toluene-d8	8.71	98	484046	53.50	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.00%	
62) 4-Bromofluorobenzene	11.14	95	171738	50.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.04%	
Target Compounds						
16) Acetone	2.45	43	30065	17.004	ug/l	96
73) Isopropylbenzene	11.02	105	24523	2.145	ug/l	99
78) n-propylbenzene	11.36	91	23074	1.754	ug/l	98
83) tert-Butylbenzene	11.77	119	7914	0.798	ug/l	91
85) sec-Butylbenzene	11.94	105	105884	9.026	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101618\
Data File : VX005361.D
Acq On : 16 Oct 2018 11:00
Operator : JC/MD
Sample : J5521-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TWP-01

Quant Time: Oct 17 05:36:40 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration

