

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101618\
 Data File : VX005364.D
 Acq On : 16 Oct 2018 12:20
 Operator : JC/MD
 Sample : J5519-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BS-DRUM

Manual Integrations
 APPROVED

MMDadoda
 10/17/2018 9:34:46 AM

Quant Time: Oct 17 06:01:11 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	302534	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	430535	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	385098	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	199551	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	166723	49.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.30%	
35) Dibromofluoromethane	5.50	113	134767	47.26	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.52%	
50) Toluene-d8	8.71	98	515727	52.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.36%	
62) 4-Bromofluorobenzene	11.14	95	175240	47.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.30%	
Target Compounds						
16) Acetone	2.45	43	7170	3.693	ug/l	100
80) 1,3,5-Trimethylbenzene	11.51	105	25136	2.474	ug/l	97
84) 1,2,4-Trimethylbenzene	11.81	105	31349	2.999	ug/l	97
86) p-Isopropyltoluene	12.07	119	2613m	0.237	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101618\
Data File : VX005364.D
Acq On : 16 Oct 2018 12:20
Operator : JC/MD
Sample : J5519-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BS-DRUM

Manual Integrations
APPROVED

MMDadoda
10/17/2018 9:34:46 AM

Quant Time: Oct 17 06:01:11 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration

