

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101618\
 Data File : VX005374.D
 Acq On : 16 Oct 2018 17:47
 Operator : JC/MD
 Sample : J5538-02 200X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TKABCD

Quant Time: Oct 17 07:14:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	287041	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	416668	50.00	ug/l	-0.01
63) Chlorobenzene-d5	10.12	117	363325	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	177814	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	161088	50.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.14%	
35) Dibromofluoromethane	5.50	113	128007	46.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.76%	
50) Toluene-d8	8.71	98	483467	51.03	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.06%	
62) 4-Bromofluorobenzene	11.14	95	159697	44.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.74%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101618\
Data File : VX005374.D
Acq On : 16 Oct 2018 17:47
Operator : JC/MD
Sample : J5538-02 200X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TKABCD

Quant Time: Oct 17 07:14:27 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration

