

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX101618\
 Data File : VX005355.D
 Acq On : 16 Oct 2018 08:08
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 17 05:01:40 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	0.391	0.410	-4.9	96	0.00
3 P	Chloromethane	0.566	0.531	6.2	96	0.00
4 C	Vinyl Chloride	0.574	0.527	8.2#	98	0.00
5 T	Bromomethane	0.356	0.320	10.1	100	0.00
6 T	Chloroethane	0.342	0.325	5.0	95	0.00
7 T	Trichlorofluoromethane	0.766	0.748	2.3	101	0.00
8 T	Diethyl Ether	0.327	0.301	8.0	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.442	0.451	-2.0	103	0.00
10 T	Methyl Iodide	0.440	0.525	-19.3	107	0.00
11 T	Tert butyl alcohol	0.157	0.141	10.2	95	-0.01
12 CM	1,1-Dichloroethene	0.438	0.413	5.7#	98	0.00
13 T	Acrolein	0.105	0.062	41.0#	61	0.00
14 T	Allyl chloride	0.961	0.899	6.5	95	0.00
15 T	Acrylonitrile	0.355	0.313	11.8	97	0.00
16 T	Acetone	0.321	0.308	4.0	100	0.00
17 T	Carbon Disulfide	1.327	1.205	9.2	98	0.00
18 T	Methyl Acetate	0.913	0.838	8.2	96	0.00
19 T	Methyl tert-butyl Ether	1.542	1.482	3.9	98	0.00
20 T	Methylene Chloride	0.520	0.475	8.7	97	0.00
21 T	trans-1,2-Dichloroethene	0.485	0.446	8.0	100	0.00
22 T	Diisopropyl ether	1.615	1.626	-0.7	100	0.00
23 T	Vinyl Acetate	1.411	1.439	-2.0	105	0.00
24 P	1,1-Dichloroethane	0.926	0.905	2.3	106	0.00
25 T	2-Butanone	0.469	0.436	7.0	93	-0.02
26 T	2,2-Dichloropropane	0.692	0.686	0.9	108	0.00
27 T	cis-1,2-Dichloroethene	0.527	0.488	7.4	100	0.00
28 T	Bromochloromethane	0.422	0.382	9.5	93	0.00
29 T	Tetrahydrofuran	0.299	0.259	13.4	84	0.00
30 C	Chloroform	0.859	0.801	6.8#	97	0.00
31 T	Cyclohexane	0.763	0.713	6.6	92	0.00
32 T	1,1,1-Trichloroethane	0.747	0.706	5.5	98	0.00
33 S	1,2-Dichloroethane-d4	0.555	0.517	6.8	101	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	98	0.00
35 S	Dibromofluoromethane	0.331	0.303	8.5	95	0.00
36 T	1,1-Dichloropropene	0.484	0.455	6.0	96	0.00
37 T	Ethyl Acetate	0.623	0.578	7.2	96	0.00
38 T	Carbon Tetrachloride	0.487	0.464	4.7	98	0.00
39 T	Methylcyclohexane	0.524	0.532	-1.5	102	0.00
40 TM	Benzene	1.453	1.322	9.0	90	0.00
41 T	Methacrylonitrile	0.334	0.296	11.4	90	0.00
42 TM	1,2-Dichloroethane	0.513	0.470	8.4	95	0.00
43 T	Isopropyl Acetate	0.830	0.813	2.0	98	-0.02
44 TM	Trichloroethene	0.416	0.367	11.8	102	0.00
45 C	1,2-Dichloropropane	0.356	0.349	2.0#	102	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX101618\
 Data File : VX005355.D
 Acq On : 16 Oct 2018 08:08
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 17 05:01:40 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.235	0.228	3.0	98	0.00
47 T	Bromodichloromethane	0.435	0.443	-1.8	105	0.00
48 T	Methyl methacrylate	0.430	0.429	0.2	97	0.00
49 T	1,4-Dioxane	0.010	0.009	10.0	94	0.00
50 S	Toluene-d8	1.137	1.170	-2.9	105	0.00
51 T	4-Methyl-2-Pentanone	0.574	0.573	0.2	100	0.00
52 CM	Toluene	0.795	0.830	-4.4#	104	0.00
53 T	t-1,3-Dichloropropene	0.471	0.509	-8.1	106	0.00
54 T	cis-1,3-Dichloropropene	0.510	0.545	-6.9	106	0.00
55 T	1,1,2-Trichloroethane	0.337	0.341	-1.2	105	0.00
56 T	Ethyl methacrylate	0.491	0.524	-6.7	103	0.00
57 T	1,3-Dichloropropane	0.562	0.570	-1.4	106	0.00
58 T	2-Chloroethyl Vinyl ether	0.275	0.283	-2.9	100	0.00
59 T	2-Hexanone	0.462	0.469	-1.5	103	0.00
60 T	Dibromochloromethane	0.348	0.375	-7.8	108	0.00
61 T	1,2-Dibromoethane	0.353	0.356	-0.8	104	0.00
62 S	4-Bromofluorobenzene	0.427	0.442	-3.5	108	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	104	0.00
64 T	Tetrachloroethene	0.403	0.421	-4.5	112	0.00
65 PM	Chlorobenzene	1.064	1.041	2.2	106	0.00
66 T	1,1,1,2-Tetrachloroethane	0.379	0.381	-0.5	105	0.00
67 C	Ethyl Benzene	1.729	1.786	-3.3#	107	0.00
68 T	m/p-Xylenes	0.671	0.704	-4.9	107	0.00
69 T	o-Xylene	0.647	0.664	-2.6	106	0.00
70 T	Styrene	1.065	1.129	-6.0	106	0.00
71 P	Bromoform	0.339	0.349	-2.9	110	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
73 T	Isopropylbenzene	2.972	3.103	-4.4	107	0.00
74 T	N-amyl acetate	1.434	1.412	1.5	104	0.00
75 P	1,1,2,2-Tetrachloroethane	1.121	1.031	8.0	105	0.00
76 T	1,2,3-Trichloropropane	0.970	0.989	-2.0	112	0.00
77 T	Bromobenzene	0.850	0.836	1.6	108	0.00
78 T	n-propylbenzene	3.420	3.585	-4.8	109	0.00
79 T	2-Chlorotoluene	2.095	2.117	-1.1	108	0.00
80 T	1,3,5-Trimethylbenzene	2.545	2.672	-5.0	107	0.00
81 T	trans-1,4-Dichloro-2-butene	0.319	0.319	0.0	108	0.00
82 T	4-Chlorotoluene	2.483	2.527	-1.8	108	0.00
83 T	tert-Butylbenzene	2.578	2.657	-3.1	108	0.00
84 T	1,2,4-Trimethylbenzene	2.619	2.761	-5.4	108	0.00
85 T	sec-Butylbenzene	3.050	3.207	-5.1	109	0.00
86 T	p-Isopropyltoluene	2.757	2.981	-8.1	111	0.00
87 T	1,3-Dichlorobenzene	1.588	1.599	-0.7	111	0.00
88 T	1,4-Dichlorobenzene	1.625	1.622	0.2	111	0.00
89 T	n-Butylbenzene	2.503	2.660	-6.3	111	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX101618\
Data File : VX005355.D
Acq On : 16 Oct 2018 08:08
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Oct 17 05:01:40 2018
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.475	0.472	0.6	108	0.00
91 T	1,2-Dichlorobenzene	1.621	1.596	1.5	110	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.274	0.251	8.4	103	0.00
93 T	1,2,4-Trichlorobenzene	1.189	1.256	-5.6	112	0.00
94 T	Hexachlorobutadiene	0.628	0.658	-4.8	117	0.00
95 T	Naphthalene	3.448	3.650	-5.9	106	0.00
96 T	1,2,3-Trichlorobenzene	1.217	1.249	-2.6	109	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6