

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101618\
 Data File : VX005356.D
 Acq On : 16 Oct 2018 08:40
 Operator : JC/MD
 Sample : VX1016WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1016WBL01

Quant Time: Oct 17 05:12:15 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	289553	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	414319	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	371704	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	181371	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	166070	51.68	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.36%	
35) Dibromofluoromethane	5.50	113	134970	49.18	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.36%	
50) Toluene-d8	8.71	98	497707	52.83	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.66%	
62) 4-Bromofluorobenzene	11.14	95	161021	45.50	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.00%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101618\
Data File : VX005356.D
Acq On : 16 Oct 2018 08:40
Operator : JC/MD
Sample : VX1016WBL01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
VX1016WBL01

Quant Time: Oct 17 05:12:15 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. -0.00 min

Lab File: VX005356.D

Acq: 16 Oct 2018 08:40

Instrument :

MSVOA_X

ClientSampleId :


VX1016WBL01

Tot Ion:168 Resp: 289553

Ion Ratio Lower Upper

168 100

99 48.4 39.9 59.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

120000

100000

80000



60000

40000

20000

0

Time--> 5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 51.677 ug/l

RT: 6.07 min Scan# 806

Delta R.T. -0.01 min

Lab File: VX005356.D


Acq: 16 Oct 2018 08:40

Tgt Ion: 65 Resp: 166070

Ion Ratio Lower Upper

65 100

67 52.7 0.0 106.8



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0


60000


40000

20000

0

Time--> 5.90 6.00 6.10 6.20 6.30









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.86 min Scan# 936
Delta R.T. -0.01 min
Lab File: VX005356.D
Acq: 16 Oct 2018 08:40

Instrument :
MSVOA_X
ClientSampleId :
VX1016WBL01

Tot Ion:114 Resp: 414319
Ion Ratio Lower Upper
114 100
63 19.7 0.0 39.0
88 15.1 0.0 30.4




Abundance Ion 113.90 (113.60 to 114.60): V
Ion 63.00 (62.70 to 63.70): VX0
Ion 87.90 (87.60 to 88.60): VX0






#35
Dibromofluoromethane
Concen: 49.185 ug/l
RT: 5.50 min Scan# 713
Delta R.T. -0.00 min
Lab File: VX005356.D
Acq: 16 Oct 2018 08:40

Tgt Ion:113 Resp: 134970
Ion Ratio Lower Upper
113 100
111 101.2 82.4 123.6
192 24.6 19.4 29.2



Abundance Ion 112.80 (112.50 to 113.50): V
Ion 110.80 (110.50 to 111.50): V
Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 52.831 ug/l
RT: 8.71 min Scan# 1240
Delta R.T. -0.00 min
Lab File: VX005356.D
Acq: 16 Oct 2018 08:40

Instrument :
MSVOA_X
ClientSampleId :
VX1016WBL01

Tot Ion: 98 Resp: 497707
Ion Ratio Lower Upper
98 100
100 63.5 52.9 79.3




Abundance Ion 98.00 (97.70 to 98.70): VX0
Ion 100.00 (99.70 to 100.70): V






#62
4-Bromofluorobenzene
Concen: 45.498 ug/l
RT: 11.14 min Scan# 1638
Delta R.T. -0.00 min
Lab File: VX005356.D
Acq: 16 Oct 2018 08:40

Tgt Ion: 95 Resp: 161021
Ion Ratio Lower Upper
95 100
174 97.7 0.0 185.2
176 93.3 0.0 178.6



Abundance Ion 94.90 (94.60 to 95.60): VX0
Ion 173.80 (173.50 to 174.50): V
Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.12 min Scan# 1470
Delta R.T. -0.00 min
Lab File: VX005356.D
Acq: 16 Oct 2018 08:40

Instrument :
MSVOA_X
ClientSampleId :
VX1016WBL01

Tot Ion:117 Resp: 371704
Ion Ratio Lower Upper
117 100
82 52.8 47.8 71.6
119 31.9 29.3 43.9




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. -0.00 min
Lab File: VX005356.D
Acq: 16 Oct 2018 08:40

Tgt Ion:152 Resp: 181371
Ion Ratio Lower Upper
152 100
115 54.4 39.0 117.0
150 156.0 0.0 351.0



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

