


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX101618\
Data File : VX005354.D
Acq On : 16 Oct 2018 07:35
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Title : SW846 8260
Last Update : Fri Oct 12 11:28:43 2018



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1630

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.9	7369	PASS
75	95	30	60	49.4	20329	PASS
95	95	100	100	100.0	41173	PASS
96	95	5	9	7.0	2883	PASS
173	174	0.00	2	1.5	563	PASS
174	95	50	100	92.7	38168	PASS
175	174	5	9	7.4	2843	PASS
176	174	95	101	95.4	36424	PASS
177	176	5	9	6.7	2455	PASS