

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101621\
 Data File : VX024837.D
 Acq On : 15 Oct 2021 18:55
 Operator : JC/MD
 Sample : M4251-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SW-RR-01-(2.0)

Quant Time: Oct 18 01:44:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 14 03:51:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	140474	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	231365	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	209714	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91987	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	104101	60.294	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 120.580%#	
35) Dibromofluoromethane	5.397	113	77031	55.927	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 111.860%	
50) Toluene-d8	8.653	98	276086	54.189	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 108.380%	
62) 4-Bromofluorobenzene	11.085	95	94453	47.136	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 94.280%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101621\
Data File : VX024837.D
Acq On : 15 Oct 2021 18:55
Operator : JC/MD
Sample : M4251-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SW-RR-01-(2.0)

Quant Time: Oct 18 01:44:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101321W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 14 03:51:38 2021
Response via : Initial Calibration

