

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101621\
 Data File : VX024838.D
 Acq On : 15 Oct 2021 19:18
 Operator : JC/MD
 Sample : M4251-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RR-54

Quant Time: Oct 18 01:44:53 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 14 03:51:38 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	137209	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	229857	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	207328	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	92650	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	101719	60.316	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 120.640%#	
35) Dibromofluoromethane	5.397	113	76153	55.652	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 111.300%	
50) Toluene-d8	8.653	98	277706	54.864	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 109.720%	
62) 4-Bromofluorobenzene	11.085	95	95492	47.967	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.940%	
Target Compounds						
					Qvalue	
30) Chloroform	5.092	83	1277	0.473	ug/l	92
52) Toluene	8.732	92	1451	0.384	ug/l	87
65) Chlorobenzene	10.085	112	1172	0.289	ug/l #	82
95) Naphthalene	13.780	128	2718	0.669	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101621\
Data File : VX024838.D
Acq On : 15 Oct 2021 19:18
Operator : JC/MD
Sample : M4251-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RR-54

Quant Time: Oct 18 01:44:53 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101321W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 14 03:51:38 2021
Response via : Initial Calibration

