

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101621\
 Data File : VX024839.D
 Acq On : 15 Oct 2021 19:42
 Operator : JC/MD
 Sample : M4251-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB

Quant Time: Oct 18 01:45:02 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 14 03:51:38 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	143621	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	236785	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	217675	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94721	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	105334	59.671	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	119.340%#
35) Dibromofluoromethane	5.397	113	78953	56.010	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	112.020%
50) Toluene-d8	8.653	98	282360	54.152	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	108.300%
62) 4-Bromofluorobenzene	11.085	95	97793	47.686	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101621\
Data File : VX024839.D
Acq On : 15 Oct 2021 19:42
Operator : JC/MD
Sample : M4251-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB

Quant Time: Oct 18 01:45:02 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101321W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 14 03:51:38 2021
Response via : Initial Calibration

