

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX101621\
 Data File : VX024840.D
 Acq On : 15 Oct 2021 20:52
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 18 01:45:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 14 03:51:38 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	50.000	53.658	-7.3	100	0.00
3 P	Chloromethane	50.000	55.191	-10.4	107	0.00
4 C	Vinyl Chloride	50.000	55.455	-10.9#	105	0.00
5 T	Bromomethane	50.000	57.914	-15.8	100	0.00
6 T	Chloroethane	50.000	52.979	-6.0	98	0.00
7 T	Trichlorofluoromethane	50.000	51.280	-2.6	96	0.00
8 T	Diethyl Ether	50.000	52.679	-5.4	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	49.335	1.3	96	0.00
10 T	Methyl Iodide	50.000	54.359	-8.7	104	0.00
11 T	Tert butyl alcohol	250.000	292.083	-16.8	106	0.00
12 CM	1,1-Dichloroethene	50.000	50.916	-1.8#	96	0.00
13 T	Acrolein	250.000	276.469	-10.6	110	0.00
14 T	Allyl chloride	50.000	55.183	-10.4	100	0.00
15 T	Acrylonitrile	250.000	289.740	-15.9	105	0.00
16 T	Acetone	250.000	247.259	1.1	85	0.00
17 T	Carbon Disulfide	50.000	48.670	2.7	96	0.00
18 T	Methyl Acetate	50.000	57.110	-14.2	106	0.00
19 T	Methyl tert-butyl Ether	50.000	56.517	-13.0	104	0.00
20 T	Methylene Chloride	50.000	53.176	-6.4	105	0.00
21 T	trans-1,2-Dichloroethene	50.000	52.555	-5.1	101	0.00
22 T	Diisopropyl ether	50.000	56.558	-13.1	104	0.00
23 T	Vinyl Acetate	250.000	287.886	-15.2	102	0.00
24 P	1,1-Dichloroethane	50.000	55.009	-10.0	104	0.00
25 T	2-Butanone	250.000	281.813	-12.7	100	0.00
26 T	2,2-Dichloropropane	50.000	49.056	1.9	90	0.00
27 T	cis-1,2-Dichloroethene	50.000	56.348	-12.7	106	0.00
28 T	Bromochloromethane	50.000	55.390	-10.8	108	0.00
29 T	Tetrahydrofuran	250.000	294.276	-17.7	107	0.00
30 C	Chloroform	50.000	55.627	-11.3#	103	0.00
31 T	Cyclohexane	50.000	54.960	-9.9	101	0.00
32 T	1,1,1-Trichloroethane	50.000	56.179	-12.4	103	0.00
33 S	1,2-Dichloroethane-d4	50.000	52.977	-6.0	104	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	98	0.00
35 S	Dibromofluoromethane	50.000	53.602	-7.2	108	0.00
36 T	1,1-Dichloropropene	50.000	52.968	-5.9	99	0.00
37 T	Ethyl Acetate	50.000	56.674	-13.3	105	0.00
38 T	Carbon Tetrachloride	50.000	54.411	-8.8	102	0.00
39 T	Methylcyclohexane	50.000	54.264	-8.5	103	0.00
40 TM	Benzene	50.000	55.310	-10.6	104	0.00
41 T	Methacrylonitrile	50.000	56.870	-13.7	104	0.00
42 TM	1,2-Dichloroethane	50.000	53.169	-6.3	100	0.00
43 T	Isopropyl Acetate	50.000	56.553	-13.1	103	0.00
44 TM	Trichloroethene	50.000	54.937	-9.9	107	0.00
45 C	1,2-Dichloropropane	50.000	57.270	-14.5#	108	0.00
46 T	Dibromomethane	50.000	56.019	-12.0	105	0.00
47 T	Bromodichloromethane	50.000	56.425	-12.8	104	0.00
48 T	Methyl methacrylate	50.000	59.175	-18.3	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101621\
 Data File : VX024840.D
 Acq On : 15 Oct 2021 20:52
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 18 01:45:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101321W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 14 03:51:38 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

49 T	1,4-Dioxane	1000.000	1259.721	-26.0#	113	0.00
50 S	Toluene-d8	50.000	55.161	-10.3	110	0.00
51 T	4-Methyl-2-Pentanone	250.000	293.867	-17.5	105	0.00
52 CM	Toluene	50.000	57.217	-14.4#	106	0.00
53 T	t-1,3-Dichloropropene	50.000	57.045	-14.1	101	0.00
54 T	cis-1,3-Dichloropropene	50.000	57.081	-14.2	102	0.00
55 T	1,1,2-Trichloroethane	50.000	58.203	-16.4	108	0.00
56 T	Ethyl methacrylate	50.000	54.302	-8.6	107	0.00
57 T	1,3-Dichloropropane	50.000	57.456	-14.9	106	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	294.371	-17.7	108	0.00
59 T	2-Hexanone	250.000	278.038	-11.2	101	0.00
60 T	Dibromochloromethane	50.000	52.968	-5.9	105	0.00
61 T	1,2-Dibromoethane	50.000	57.121	-14.2	105	0.00
62 S	4-Bromofluorobenzene	50.000	54.300	-8.6	108	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	100	0.00
64 T	Tetrachloroethene	50.000	53.760	-7.5	106	0.00
65 PM	Chlorobenzene	50.000	55.002	-10.0	106	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	55.362	-10.7	105	0.00
67 C	Ethyl Benzene	50.000	55.668	-11.3#	105	0.00
68 T	m/p-Xylenes	100.000	115.305	-15.3	105	0.00
69 T	o-Xylene	50.000	57.317	-14.6	106	0.00
70 T	Styrene	50.000	53.941	-7.9	105	0.00
71 P	Bromoform	50.000	50.403	-0.8	104	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	100	0.00
73 T	Isopropylbenzene	50.000	56.071	-12.1	104	0.00
74 T	N-amyl acetate	50.000	55.614	-11.2	101	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	55.147	-10.3	106	0.00
76 T	1,2,3-Trichloropropane	50.000	55.988	-12.0	105	0.00
77 T	Bromobenzene	50.000	55.233	-10.5	107	0.00
78 T	n-propylbenzene	50.000	55.998	-12.0	103	0.00
79 T	2-Chlorotoluene	50.000	55.693	-11.4	104	0.00
80 T	1,3,5-Trimethylbenzene	50.000	57.140	-14.3	103	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	51.619	-3.2	101	0.00
82 T	4-Chlorotoluene	50.000	54.647	-9.3	102	0.00
83 T	tert-Butylbenzene	50.000	57.374	-14.7	104	0.00
84 T	1,2,4-Trimethylbenzene	50.000	56.960	-13.9	103	0.00
85 T	sec-Butylbenzene	50.000	56.579	-13.2	103	0.00
86 T	p-Isopropyltoluene	50.000	58.341	-16.7	104	0.00
87 T	1,3-Dichlorobenzene	50.000	54.002	-8.0	104	0.00
88 T	1,4-Dichlorobenzene	50.000	54.097	-8.2	106	0.00
89 T	n-Butylbenzene	50.000	55.785	-11.6	101	0.00
90 T	Hexachloroethane	50.000	56.594	-13.2	104	0.00
91 T	1,2-Dichlorobenzene	50.000	54.203	-8.4	104	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	58.293	-16.6	107	0.00
93 T	1,2,4-Trichlorobenzene	50.000	56.230	-12.5	104	0.00
94 T	Hexachlorobutadiene	50.000	53.110	-6.2	103	0.00
95 T	Naphthalene	50.000	54.622	-9.2	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101621\
Data File : VX024840.D
Acq On : 15 Oct 2021 20:52
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Oct 18 01:45:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101321W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 14 03:51:38 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	58.082	-16.2	105	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6