

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101624\
 Data File : VX043419.D
 Acq On : 16 Oct 2024 13:14
 Operator : JC/MD
 Sample : P4405-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-121

Quant Time: Oct 17 01:53:38 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	80645	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	145095	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	122380	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	52081	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	66711	50.652	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.300%
35) Dibromofluoromethane	5.385	113	47594	47.564	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.120%
50) Toluene-d8	8.647	98	172073	49.622	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.240%
62) 4-Bromofluorobenzene	11.079	95	60245	47.822	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	95.640%
Target Compounds						
						Qvalue
16) Acetone	2.392	43	1617	2.940	ug/l	92
20) Methylene Chloride	2.788	84	2514	2.479	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101624\
Data File : VX043419.D
Acq On : 16 Oct 2024 13:14
Operator : JC/MD
Sample : P4405-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-121

Quant Time: Oct 17 01:53:38 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

