

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101624\
 Data File : VX043420.D
 Acq On : 16 Oct 2024 15:12
 Operator : JC/MD
 Sample : P4417-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HOPPER-A-WATER

Quant Time: Oct 17 01:54:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	77102	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.751	114	142201	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	130823	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	59487	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	64010	50.834	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.660%
35) Dibromofluoromethane	5.379	113	47085	48.013	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.020%
50) Toluene-d8	8.641	98	171401	50.434	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	100.860%
62) 4-Bromofluorobenzene	11.079	95	65588	53.123	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	106.240%
Target Compounds						
16) Acetone	2.386	43	10002	19.021	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101624\
Data File : VX043420.D
Acq On : 16 Oct 2024 15:12
Operator : JC/MD
Sample : P4417-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
HOPPER-A-WATER

Quant Time: Oct 17 01:54:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

