

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX101624\
 Data File : VX043414.D
 Acq On : 16 Oct 2024 11:08
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 17 01:49:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	87	0.00
2 T	Dichlorodifluoromethane	50.000	48.686	2.6	82	0.00
3 P	Chloromethane	50.000	47.743	4.5	83	0.00
4 C	Vinyl Chloride	50.000	49.268	1.5#	86	0.00
5 T	Bromomethane	50.000	51.271	-2.5	92	0.00
6 T	Chloroethane	50.000	55.128	-10.3	95	0.00
7 T	Trichlorofluoromethane	50.000	54.161	-8.3	97	0.00
8 T	Diethyl Ether	50.000	50.271	-0.5	89	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	51.453	-2.9	90	0.00
10 T	Methyl Iodide	50.000	49.233	1.5	84	0.00
11 T	Tert butyl alcohol	250.000	279.033	-11.6	95	0.00
12 CM	1,1-Dichloroethene	50.000	46.676	6.6#	82	0.00
13 T	Acrolein	250.000	242.671	2.9	85	0.00
14 T	Allyl chloride	50.000	51.224	-2.4	89	0.00
15 T	Acrylonitrile	250.000	251.829	-0.7	87	0.00
16 T	Acetone	250.000	281.723	-12.7	106	0.00
17 T	Carbon Disulfide	50.000	43.278	13.4	77	0.00
18 T	Methyl Acetate	50.000	48.232	3.5	81	0.00
19 T	Methyl tert-butyl Ether	50.000	52.163	-4.3	91	0.00
20 T	Methylene Chloride	50.000	48.727	2.5	87	0.00
21 T	trans-1,2-Dichloroethene	50.000	49.400	1.2	87	0.00
22 T	Diisopropyl ether	50.000	51.878	-3.8	90	0.00
23 T	Vinyl Acetate	250.000	207.325	17.1	67	0.00
24 P	1,1-Dichloroethane	50.000	51.118	-2.2	89	0.00
25 T	2-Butanone	250.000	262.625	-5.1	91	0.00
26 T	2,2-Dichloropropane	50.000	52.669	-5.3	93	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.377	-0.8	89	0.00
28 T	Bromochloromethane	50.000	42.402	15.2	77	0.00
29 T	Tetrahydrofuran	250.000	247.793	0.9	88	0.00
30 C	Chloroform	50.000	50.661	-1.3#	90	0.00
31 T	Cyclohexane	50.000	49.687	0.6	89	0.00
32 T	1,1,1-Trichloroethane	50.000	50.011	-0.0	87	0.00
33 S	1,2-Dichloroethane-d4	50.000	46.570	6.9	82	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	86	0.00
35 S	Dibromofluoromethane	50.000	47.329	5.3	83	0.00
36 T	1,1-Dichloropropene	50.000	51.325	-2.7	89	0.00
37 T	Ethyl Acetate	50.000	50.947	-1.9	88	0.00
38 T	Carbon Tetrachloride	50.000	51.203	-2.4	86	0.00
39 T	Methylcyclohexane	50.000	52.459	-4.9	90	0.00
40 TM	Benzene	50.000	50.527	-1.1	87	0.00
41 T	Methacrylonitrile	50.000	53.383	-6.8	86	0.00
42 TM	1,2-Dichloroethane	50.000	52.580	-5.2	89	0.00
43 T	Isopropyl Acetate	50.000	52.622	-5.2	88	0.00
44 TM	Trichloroethene	50.000	49.499	1.0	87	0.00
45 C	1,2-Dichloropropane	50.000	52.052	-4.1#	89	0.00
46 T	Dibromomethane	50.000	52.455	-4.9	89	0.00
47 T	Bromodichloromethane	50.000	52.571	-5.1	88	0.00
48 T	Methyl methacrylate	50.000	52.921	-5.8	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101624\
 Data File : VX043414.D
 Acq On : 16 Oct 2024 11:08
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 17 01:49:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 02 16:50:57 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	993.312	0.7	83	0.00
50 S	Toluene-d8	50.000	45.608	8.8	77	0.00
51 T	4-Methyl-2-Pentanone	250.000	258.826	-3.5	86	0.00
52 CM	Toluene	50.000	51.852	-3.7#	87	0.00
53 T	t-1,3-Dichloropropene	50.000	54.009	-8.0	87	0.00
54 T	cis-1,3-Dichloropropene	50.000	53.235	-6.5	89	0.00
55 T	1,1,2-Trichloroethane	50.000	51.835	-3.7	86	0.00
56 T	Ethyl methacrylate	50.000	53.857	-7.7	86	0.00
57 T	1,3-Dichloropropane	50.000	52.277	-4.6	87	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	259.945	-4.0	82	0.00
59 T	2-Hexanone	250.000	263.267	-5.3	87	0.00
60 T	Dibromochloromethane	50.000	51.677	-3.4	83	0.00
61 T	1,2-Dibromoethane	50.000	51.937	-3.9	87	0.00
62 S	4-Bromofluorobenzene	50.000	48.617	2.8	80	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	84	0.00
64 T	Tetrachloroethene	50.000	50.993	-2.0	87	0.00
65 PM	Chlorobenzene	50.000	50.780	-1.6	87	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	52.472	-4.9	86	0.00
67 C	Ethyl Benzene	50.000	52.315	-4.6#	87	0.00
68 T	m/p-Xylenes	100.000	102.333	-2.3	86	0.00
69 T	o-Xylene	50.000	50.857	-1.7	86	0.00
70 T	Styrene	50.000	52.735	-5.5	86	0.00
71 P	Bromoform	50.000	50.332	-0.7	81	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	86	0.00
73 T	Isopropylbenzene	50.000	50.218	-0.4	86	0.00
74 T	N-amyl acetate	50.000	49.825	0.3	86	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	50.005	-0.0	87	0.00
76 T	1,2,3-Trichloropropane	50.000	49.512	1.0	88	0.00
77 T	Bromobenzene	50.000	48.223	3.6	84	0.00
78 T	n-propylbenzene	50.000	52.356	-4.7	87	0.00
79 T	2-Chlorotoluene	50.000	49.612	0.8	88	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.310	-2.6	89	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	49.400	1.2	87	0.00
82 T	4-Chlorotoluene	50.000	50.476	-1.0	88	0.00
83 T	tert-Butylbenzene	50.000	49.765	0.5	86	0.00
84 T	1,2,4-Trimethylbenzene	50.000	50.750	-1.5	87	0.00
85 T	sec-Butylbenzene	50.000	51.650	-3.3	88	0.00
86 T	p-Isopropyltoluene	50.000	51.844	-3.7	88	0.00
87 T	1,3-Dichlorobenzene	50.000	50.634	-1.3	89	0.00
88 T	1,4-Dichlorobenzene	50.000	48.963	2.1	88	0.00
89 T	n-Butylbenzene	50.000	54.248	-8.5	90	0.00
90 T	Hexachloroethane	50.000	50.142	-0.3	84	0.00
91 T	1,2-Dichlorobenzene	50.000	49.241	1.5	87	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	48.797	2.4	83	0.00
93 T	1,2,4-Trichlorobenzene	50.000	52.731	-5.5	89	0.00
94 T	Hexachlorobutadiene	50.000	53.577	-7.2	89	0.00
95 T	Naphthalene	50.000	50.093	-0.2	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101624\
Data File : VX043414.D
Acq On : 16 Oct 2024 11:08
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Oct 17 01:49:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100124W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 02 16:50:57 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	51.204	-2.4	88	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6