

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101718\
 Data File : VX005396.D
 Acq On : 17 Oct 2018 17:35
 Operator : JC/MD
 Sample : VSTDCCC050EC
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Oct 18 11:12:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	291210	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	450347	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	386246	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	218923	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	143522	44.41	ug/l	0.00
Spiked Amount			Recovery	=	88.82%	
35) Dibromofluoromethane	5.50	113	121105	40.60	ug/l	0.00
Spiked Amount			Recovery	=	81.20%	
50) Toluene-d8	8.71	98	460385	44.96	ug/l	0.00
Spiked Amount			Recovery	=	89.92%	
62) 4-Bromofluorobenzene	11.14	95	175930	45.73	ug/l	0.00
Spiked Amount			Recovery	=	91.46%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) Dichlorodifluoromethane	1.20	85	103359	45.395	ug/l	96
3) Chloromethane	1.32	50	142806	43.285	ug/l	100
4) Vinyl Chloride	1.41	62	154845	46.331	ug/l	91
5) Bromomethane	1.64	94	91531	48.509	ug/l	94
6) Chloroethane	1.71	64	93813	47.280	ug/l	97
7) Trichlorofluoromethane	1.92	101	212693	47.689	ug/l	97
8) Diethyl Ether	2.18	74	97060	50.962	ug/l	90
9) 1,1,2-Trichlorotrifluoroet	2.38	101	134865	52.391	ug/l	96
10) Methyl Iodide	2.51	142	152313	53.765	ug/l	99
11) Tert butyl alcohol	3.08	59	222810	244.392	ug/l	99
12) 1,1-Dichloroethene	2.37	96	138309	54.228	ug/l	87
13) Acrolein	2.29	56	167519	288.278	ug/l	97
14) Allyl chloride	2.72	41	270141	48.242	ug/l	95
15) Acrylonitrile	3.15	53	497160	240.643	ug/l	100
16) Acetone	2.45	43	466113	249.432	ug/l	93
17) Carbon Disulfide	2.57	76	333424	43.154	ug/l	99
18) Methyl Acetate	2.78	43	265552	49.950	ug/l	95
19) Methyl tert-butyl Ether	3.20	73	463382	51.586	ug/l	98
20) Methylene Chloride	2.85	84	142218	52.292	ug/l	99
21) trans-1,2-Dichloroethene	3.16	96	133114	47.095	ug/l	99
22) Diisopropyl ether	3.87	45	444506	47.261	ug/l	92
23) Vinyl Acetate	3.82	43	1956280	237.971	ug/l	97
24) 1,1-Dichloroethane	3.69	63	254389	47.193	ug/l	99
25) 2-Butanone	4.70	43	653859	239.148	ug/l	96
26) 2,2-Dichloropropane	4.58	77	200961	49.849	ug/l	99
27) cis-1,2-Dichloroethene	4.60	96	153215	49.878	ug/l	99
28) Bromochloromethane	5.02	49	116353	47.350	ug/l	96
29) Tetrahydrofuran	5.15	42	412373	236.450	ug/l	96
30) Chloroform	5.21	83	237650	47.499	ug/l	100
31) Cyclohexane	5.57	56	208897	47.019	ug/l	95
32) 1,1,1-Trichloroethane	5.49	97	206137	47.350	ug/l	99
36) 1,1-Dichloropropene	5.79	75	178275	40.874	ug/l	99
37) Ethyl Acetate	4.85	43	232215	41.399	ug/l	98
38) Carbon Tetrachloride	5.78	117	182601	41.666	ug/l	98

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101718\
 Data File : VX005396.D
 Acq On : 17 Oct 2018 17:35
 Operator : JC/MD
 Sample : VSTDCCC050EC
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleID :
 VSTDCCC050EC

Quant Time: Oct 18 11:12:10 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	7.46	83	219012	46.396	ug/l	96
40) Benzene	6.13	78	601552	45.977	ug/l	98
41) Methacrylonitrile	5.06	41	127498	42.399	ug/l	97
42) 1,2-Dichloroethane	6.20	62	188622	40.828	ug/l	99
43) Isopropyl Acetate	6.46	43	372008	49.788	ug/l	97
44) Trichloroethene	7.21	130	150990	45.136	ug/l	97
45) 1,2-Dichloropropane	7.52	63	149220	46.494	ug/l	89
46) Dibromomethane	7.66	93	92851	43.833	ug/l	98
47) Bromodichloromethane	7.90	83	178750	45.621	ug/l	99
48) Methyl methacrylate	7.77	41	180796	46.654	ug/l	96
49) 1,4-Dioxane	7.76	88	79815	892.959	ug/l	97
51) 4-Methyl-2-Pentanone	8.65	43	1225868	236.946	ug/l	97
52) Toluene	8.79	92	340047	47.506	ug/l	97
53) t-1,3-Dichloropropene	9.04	75	210592	49.664	ug/l	100
54) cis-1,3-Dichloropropene	8.44	75	225422	49.063	ug/l	97
55) 1,1,2-Trichloroethane	9.22	97	142469	46.991	ug/l	97
56) Ethyl methacrylate	9.18	69	232380	52.577	ug/l	96
57) 1,3-Dichloropropane	9.37	76	235662	46.570	ug/l	100
58) 2-Chloroethyl Vinyl ether	8.32	63	613634	247.561	ug/l	95
59) 2-Hexanone	9.50	43	977200	234.809	ug/l	96
60) Dibromochloromethane	9.59	129	149733	47.778	ug/l	99
61) 1,2-Dibromoethane	9.67	107	149266	46.912	ug/l	99
64) Tetrachloroethene	9.34	164	150150	48.288	ug/l	98
65) Chlorobenzene	10.14	112	391122	47.605	ug/l	94
66) 1,1,1,2-Tetrachloroethane	10.22	131	143583	49.041	ug/l	99
67) Ethyl Benzene	10.25	91	666126	49.873	ug/l	96
68) m/p-Xylenes	10.36	106	518759	100.068	ug/l	96
69) o-Xylene	10.70	106	254047	50.812	ug/l	98
70) Styrene	10.71	104	424701	51.645	ug/l	98
71) Bromoform	10.86	173	128692	49.187	ug/l	99
73) Isopropylbenzene	11.02	105	687251	52.807	ug/l	100
74) N-amyl acetate	10.90	43	331992	52.890	ug/l	94
75) 1,1,2,2-Tetrachloroethane	11.27	83	234104	47.685	ug/l	99
76) 1,2,3-Trichloropropane	11.30	75	276209	65.044	ug/l	88
77) Bromobenzene	11.26	156	185786	49.927	ug/l	99
78) n-propylbenzene	11.36	91	782902	52.277	ug/l	100
79) 2-Chlorotoluene	11.42	91	463826	50.570	ug/l	99
80) 1,3,5-Trimethylbenzene	11.51	105	587607	52.725	ug/l	97
81) trans-1,4-Dichloro-2-buten	11.07	75	71119	50.888	ug/l	95
82) 4-Chlorotoluene	11.51	91	547574	50.363	ug/l	100
83) tert-Butylbenzene	11.77	119	597446	52.938	ug/l	98
84) 1,2,4-Trimethylbenzene	11.80	105	598503	52.184	ug/l	99
85) sec-Butylbenzene	11.94	105	718542	53.800	ug/l	99
86) p-Isopropyltoluene	12.07	119	644836	53.418	ug/l	100
87) 1,3-Dichlorobenzene	12.02	146	345930	49.750	ug/l	100
88) 1,4-Dichlorobenzene	12.10	146	350057	49.211	ug/l	99
89) n-Butylbenzene	12.39	91	577172	52.659	ug/l	98
90) Hexachloroethane	12.60	117	103019	49.519	ug/l	95
91) 1,2-Dichlorobenzene	12.39	146	346597	48.846	ug/l	99
92) 1,2-Dibromo-3-Chloropropan	13.00	75	58181	48.584	ug/l	90


Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101718\
Data File : VX005396.D
Acq On : 17 Oct 2018 17:35
Operator : JC/MD
Sample : VSTDCCC050EC
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Quant Time: Oct 18 11:12:10 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
93) 1,2,4-Trichlorobenzene	13.65	180	279721	53.731	ug/l	98
94) Hexachlorobutadiene	13.79	225	136068	49.488	ug/l	97
95) Naphthalene	13.83	128	849462	56.264	ug/l	100
96) 1,2,3-Trichlorobenzene	14.02	180	279469	52.466	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.67 min Scan# 740

Delta R.T. -0.00 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050EC

Tot Ion:168 Resp: 291210

Ion Ratio Lower Upper




168 100

99 50.7 39.9 59.9



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0



#2

Dichlorodifluoromethane

Concen: 45.395 ug/l

RT: 1.20 min Scan# 7

Delta R.T. -0.00 min

Lab File: VX005396.D


Acq: 17 Oct 2018 17:35

Tgt Ion: 85 Resp: 103359

Ion Ratio Lower Upper



85 100


87 31.6 17.0 50.9



Abundance Ion 84.90 (84.60 to 85.60): VX0

Ion 86.90 (86.60 to 87.60): VX0





#3

Chloromethane

Concen: 43.285 ug/l

RT: 1.32 min Scan# 28

Delta R.T. -0.00 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tot Ion: 50 Resp: 142806

Ion Ratio Lower Upper

50 100

52 32.7 26.2 39.2



#4

Vinyl Chloride

Concen: 46.331 ug/l

RT: 1.41 min Scan# 42

Delta R.T. 0.01 min

Lab File: VX005396.D




Acq: 17 Oct 2018 17:35


Tgt Ion: 62 Resp: 154845

Ion Ratio Lower Upper

62 100

64 27.4 25.8 38.8





#5

Bromomethane

Concen: 48.509 ug/l

RT: 1.64 min Scan# 79

Delta R.T. 0.00 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050EC

Tot Ion: 94 Resp: 91531

Ion Ratio Lower Upper

94 100

96 87.0 74.5 111.7



Abundance Ion 93.90 (93.60 to 94.60): VX0

Ion 95.90 (95.60 to 96.60): VX0

1.64

100000

80000

60000



40000

20000

0

Time-->

1.55 1.60 1.65 1.70



#6

Chloroethane

Concen: 47.280 ug/l

RT: 1.71 min Scan# 91

Delta R.T. 0.00 min

Lab File: VX005396.D


Acq: 17 Oct 2018 17:35

Tgt Ion: 64 Resp: 93813

Ion Ratio Lower Upper

64 100

66 29.3 24.6 36.8



Abundance Ion 63.90 (63.60 to 64.60): VX0

Ion 65.90 (65.60 to 66.60): VX0

80000

60000


40000


20000

0

Time-->

1.65 1.70 1.75






#7




Trichlorofluoromethane
Concen: 47.689 ug/l
RT: 1.92 min Scan# 126
Delta R.T. -0.00 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion:101 Resp: 212693
Ion Ratio Lower Upper
101 100
103 63.1 48.9 73.3




Abundance Ion 100.90 (100.60 to 101.60): V
Ion 102.80 (102.50 to 103.50): V





#8


Diethyl Ether
Concen: 50.962 ug/l
RT: 2.18 min Scan# 169
Delta R.T. -0.01 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35

Tgt Ion: 74 Resp: 97060
Ion Ratio Lower Upper
74 100
45 106.2 48.3 144.8



Abundance Ion 74.00 (73.70 to 74.70): VX0
Ion 45.00 (44.70 to 45.70): VX0





#9

1,1,2-Trichlorotrifluoroethane

Concen: 52.391 ug/l

RT: 2.38 min Scan# 201

Delta R.T. -0.00 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC


Tot Ion:101 Resp: 134865

Ion Ratio Lower Upper

101 100

85 40.7 34.2 51.4




151 84.9 65.0 97.6



Abundance Ion 100.90 (100.60 to 101.60): V

Ion 84.90 (84.60 to 85.60): VX0

Ion 150.80 (150.50 to 151.50): V



#10

Methyl Iodide

Concen: 53.765 ug/l

RT: 2.51 min Scan# 222

Delta R.T. 0.00 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35


Tgt Ion:142 Resp: 152313

Ion Ratio Lower Upper

142 100

127 41.2 33.8 50.6



141 14.4 11.4 17.0




Abundance Ion 141.80 (141.50 to 142.50): V

Ion 126.80 (126.50 to 127.50): V

Ion 140.80 (140.50 to 141.50): V









#11

Tert butyl alcohol
 Concen: 244.392 ug/l
 RT: 3.08 min Scan# 316
 Delta R.T. -0.00 min
 Lab File: VX005396.D
 Acq: 17 Oct 2018 17:35

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC




Tot Ion: 59 Resp: 222810
 Ion Ratio Lower Upper
 59 100
 57 10.0 8.2 12.4




#12

1,1-Dichloroethene
 Concen: 54.228 ug/l
 RT: 2.37 min Scan# 199
 Delta R.T. -0.01 min
 Lab File: VX005396.D
 Acq: 17 Oct 2018 17:35

Tgt Ion: 96 Resp: 138309
 Ion Ratio Lower Upper
 96 100
 61 142.8 128.8 193.2
 98 56.5 52.2 78.2





#13

Acrolein

Concen: 288.278 ug/l

RT: 2.29 min Scan# 187

Delta R.T. 0.00 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tot Ion: 56 Resp: 167519

Ion Ratio Lower Upper

56 100

55 70.8 54.7 82.1



#14

Allyl chloride

Concen: 48.242 ug/l

RT: 2.72 min Scan# 257

Delta R.T. -0.00 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35




Tgt Ion: 41 Resp: 270141


Ion Ratio Lower Upper

41 100

39 58.1 48.9 73.3

76 30.1 26.7 40.1









#15
 Acrylonitrile
 Concen: 240.643 ug/l
 RT: 3.15 min Scan# 327
 Delta R.T. -0.00 min
 Lab File: VX005396.D
 Acq: 17 Oct 2018 17:35




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tot Ion: 53 Resp: 497160
 Ion Ratio Lower Upper
 53 100
 52 81.7 65.2 97.8
 51 34.7 28.2 42.2



#16
 Acetone
 Concen: 249.432 ug/l
 RT: 2.45 min Scan# 212
 Delta R.T. -0.00 min
 Lab File: VX005396.D
 Acq: 17 Oct 2018 17:35

Tgt Ion: 43 Resp: 466113
 Ion Ratio Lower Upper
 43 100
 58 29.1 26.2 39.4





#17

Carbon Disulfide

Concen: 43.154 ug/l

RT: 2.57 min Scan# 232

Delta R.T. -0.00 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050EC

Tot Ion: 76 Resp: 333424

Ion Ratio Lower Upper




76 100

78 9.0 7.0 10.6



Abundance Ion 75.90 (75.60 to 76.60): VX0

Ion 77.90 (77.60 to 78.60): VX0



#18

Methyl Acetate

Concen: 49.950 ug/l

RT: 2.78 min Scan# 266

Delta R.T. -0.01 min

Lab File: VX005396.D


Acq: 17 Oct 2018 17:35

Tgt Ion: 43 Resp: 265552

Ion Ratio Lower Upper



43 100


74 21.7 19.4 29.2



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 74.00 (73.70 to 74.70): VX0









#19
Methyl tert-butyl Ether
Concen: 51.586 ug/l
RT: 3.20 min Scan# 336
Delta R.T. -0.01 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tot Ion: 73 Resp: 463382
Ion Ratio Lower Upper
73 100
57 22.1 17.1 25.7



#20
Methylene Chloride
Concen: 52.292 ug/l
RT: 2.85 min Scan# 279
Delta R.T. 0.00 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35

Tgt Ion: 84 Resp: 142218
Ion Ratio Lower Upper
84 100
49 126.4 101.2 151.8
51 38.5 29.5 44.3
86 62.5 52.3 78.5









#21
trans-1,2-Dichloroethene
Concen: 47.095 ug/l
RT: 3.16 min Scan# 329
Delta R.T. -0.00 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tot Ion: 96 Resp: 133114
Ion Ratio Lower Upper
96 100
61 144.8 113.8 170.6
98 64.7 51.8 77.8



#22
Diisopropyl ether
Concen: 47.261 ug/l
RT: 3.87 min Scan# 445
Delta R.T. -0.00 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35

Tgt Ion: 45 Resp: 444506
Ion Ratio Lower Upper
45 100
43 61.7 42.8 64.2
87 26.3 22.6 34.0
59 11.4 9.6 14.4









#23
 Vinyl Acetate
 Concen: 237.971 ug/l
 RT: 3.82 min Scan# 438
 Delta R.T. 0.01 min
 Lab File: VX005396.D
 Acq: 17 Oct 2018 17:35




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tot Ion: 43 Resp: 1956280
 Ion Ratio Lower Upper
 43 100
 86 9.9 8.9 13.3



#24
 1,1-Dichloroethane
 Concen: 47.193 ug/l
 RT: 3.69 min Scan# 416
 Delta R.T. -0.00 min
 Lab File: VX005396.D
 Acq: 17 Oct 2018 17:35

Tgt Ion: 63 Resp: 254389
 Ion Ratio Lower Upper
 63 100
 98 7.0 3.5 10.4
 100 4.2 2.4 7.1





#25

2-Butanone

Concen: 239.148 ug/l

RT: 4.70 min Scan# 582

Delta R.T. -0.01 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050EC

Tot Ion: 43 Resp: 653859

Ion Ratio Lower Upper

43 100

72 25.2 22.0 33.0



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 72.00 (71.70 to 72.70): VX0

250000

200000

150000

100000

50000



0

4.60

4.70

4.80

Time-->



#26

2,2-Dichloropropane

Concen: 49.849 ug/l

RT: 4.58 min Scan# 562

Delta R.T. -0.00 min

Lab File: VX005396.D


Acq: 17 Oct 2018 17:35

Tgt Ion: 77 Resp: 200961

Ion Ratio Lower Upper

77 100

97 23.9 11.7 35.0



Abundance Ion 76.90 (76.60 to 77.60): VX0

Ion 96.90 (96.60 to 97.60): VX0

80000

60000

40000

20000

0


4.40


4.50

4.60

4.70

Time-->









#27

cis-1,2-Dichloroethene
Concen: 49.878 ug/l
RT: 4.60 min Scan# 565
Delta R.T. -0.00 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC




Tot Ion	Ratio	Lower	Upper
96	100		
61	142.9	0.0	282.6
98	64.1	0.0	130.2




#28

Bromochloromethane
Concen: 47.350 ug/l
RT: 5.02 min Scan# 634
Delta R.T. -0.00 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35

Tgt Ion	Ratio	Lower	Upper
49	100		
129	2.1	0.0	4.2
130	80.7	67.5	101.3









#29
Tetrahydrofuran
Concen: 236.450 ug/l
RT: 5.15 min Scan# 656
Delta R.T. -0.00 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion: 42 Resp: 412373
Ion Ratio Lower Upper
42 100
72 43.3 37.4 56.0
71 41.3 34.5 51.7




Abundance Ion 42.00 (41.70 to 42.70): VX0
Ion 72.00 (71.70 to 72.70): VX0
Ion 71.00 (70.70 to 71.70): VX0






#30
Chloroform
Concen: 47.499 ug/l
RT: 5.21 min Scan# 666
Delta R.T. -0.00 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35

Tgt Ion: 83 Resp: 237650
Ion Ratio Lower Upper
83 100
85 65.7 52.6 79.0



Abundance Ion 82.90 (82.60 to 83.60): VX0
Ion 84.90 (84.60 to 85.60): VX0









#31
Cyclohexane
Concen: 47.019 ug/l
RT: 5.57 min Scan# 725
Delta R.T. 0.01 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tot Ion: 56 Resp: 208897
Ion Ratio Lower Upper
56 100
69 30.5 25.4 38.2
84 84.0 71.4 107.2



#32
1,1,1-Trichloroethane
Concen: 47.350 ug/l
RT: 5.49 min Scan# 712
Delta R.T. 0.01 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35

Tgt Ion: 97 Resp: 206137
Ion Ratio Lower Upper
97 100
99 65.3 52.0 78.0
61 41.9 34.9 52.3





#33

1,2-Dichloroethane-d4

Concen: 44.407 ug/l

RT: 6.07 min Scan# 806

Delta R.T. -0.01 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050EC

Tot Ion: 65 Resp: 143522

Ion Ratio Lower Upper


65 100

67 54.0 0.0 106.8





Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0



Time-->



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.86 min Scan# 936

Delta R.T. -0.01 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35


Tgt Ion: 114 Resp: 450347

Ion Ratio Lower Upper

114 100

63 16.0 0.0 39.0


88 11.8 0.0 30.4




Abundance Ion 113.90 (113.60 to 114.60): V


Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



Time-->





#35

Dibromofluoromethane

Concen: 40.602 ug/l

RT: 5.50 min Scan# 713

Delta R.T. -0.00 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC


Tot Ion:113 Resp: 121105

Ion Ratio Lower Upper

113 100

111 102.0 82.4 123.6




192 24.4 19.4 29.2



Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V



#36

1,1-Dichloropropene

Concen: 40.874 ug/l

RT: 5.79 min Scan# 761

Delta R.T. -0.01 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35


Tgt Ion: 75 Resp: 178275

Ion Ratio Lower Upper

75 100

110 36.0 18.0 54.0



77 31.6 24.6 36.8




Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 109.90 (109.60 to 110.60): V

Ion 76.90 (76.60 to 77.60): VX0









#37
Ethyl Acetate
Concen: 41.399 ug/l
RT: 4.85 min Scan# 607
Delta R.T. -0.01 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tot Ion: 43 Resp: 232215
Ion Ratio Lower Upper
43 100
61 13.7 11.5 17.3
70 10.4 9.0 13.6



#38
Carbon Tetrachloride
Concen: 41.666 ug/l
RT: 5.78 min Scan# 759
Delta R.T. -0.00 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35

Tgt Ion: 117 Resp: 182601
Ion Ratio Lower Upper
117 100
119 96.0 78.8 118.2
121 31.5 24.9 37.3









#39
Methvlcvclohexane
Concen: 46.396 ug/l
RT: 7.46 min Scan# 1035
Delta R.T. 0.01 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion	Ratio	Lower	Upper
83	100		
55	77.6	61.0	91.4
98	44.8	39.6	59.4




Abundance Ion 83.00 (82.70 to 83.70): VX0
Ion 55.00 (54.70 to 55.70): VX0
Ion 98.00 (97.70 to 98.70): VX0






#40
Benzene
Concen: 45.977 ug/l
RT: 6.13 min Scan# 817
Delta R.T. -0.01 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35

Tgt Ion	Ratio	Lower	Upper
78	100		
77	22.8	19.0	28.4



Abundance Ion 78.00 (77.70 to 78.70): VX0
Ion 77.00 (76.70 to 77.70): VX0









#41
Methacrylonitrile
Concen: 42.399 ug/l
RT: 5.06 min Scan# 640
Delta R.T. -0.01 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion:	41	Resp:	127498
Ion	Ratio	Lower	Upper
41	100		
39	52.4	42.4	63.6
67	69.5	59.2	88.8
52	28.1	23.0	34.4




Abundance Ion 41.00 (40.70 to 41.70): VX0
Ion 39.00 (38.70 to 39.70): VX0
Ion 67.00 (66.70 to 67.70): VX0
Ion 52.00 (51.70 to 52.70): VX0






#42
1,2-Dichloroethane
Concen: 40.828 ug/l
RT: 6.20 min Scan# 827
Delta R.T. -0.00 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35

Tgt Ion:	62	Resp:	188622
Ion	Ratio	Lower	Upper
62	100		
98	9.8	0.0	20.6



Abundance Ion 61.90 (61.60 to 62.60): VX0
Ion 97.90 (97.60 to 98.60): VX0









#43

Isopropyl Acetate
 Concen: 49.788 ug/l
 RT: 6.46 min Scan# 871
 Delta R.T. -0.01 min
 Lab File: VX005396.D
 Acq: 17 Oct 2018 17:35

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC




Tot Ion: 43 Resp: 372008
 Ion Ratio Lower Upper
 43 100
 61 20.0 17.0 25.4
 87 13.1 11.4 17.2




#44

Trichloroethene
 Concen: 45.136 ug/l
 RT: 7.21 min Scan# 994
 Delta R.T. 0.01 min
 Lab File: VX005396.D
 Acq: 17 Oct 2018 17:35

Tgt Ion:130 Resp: 150990
 Ion Ratio Lower Upper
 130 100
 95 93.2 0.0 181.6









#45
 1,2-Dichloropropane
 Concen: 46.494 ug/l
 RT: 7.52 min Scan# 1044
 Delta R.T. -0.00 min
 Lab File: VX005396.D
 Acq: 17 Oct 2018 17:35




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tot Ion: 63 Resp: 149220
 Ion Ratio Lower Upper
 63 100
 65 36.3 24.3 36.5



#46
 Dibromomethane
 Concen: 43.833 ug/l
 RT: 7.66 min Scan# 1068
 Delta R.T. -0.00 min
 Lab File: VX005396.D
 Acq: 17 Oct 2018 17:35

Tgt Ion: 93 Resp: 92851
 Ion Ratio Lower Upper
 93 100
 95 83.3 65.5 98.3
 174 120.7 94.2 141.4





#47

Bromodichloromethane

Concen: 45.621 ug/l

RT: 7.90 min Scan# 1106

Delta R.T. -0.00 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC


Tot Ion: 83 Resp: 178750

Ion Ratio Lower Upper




83 100

85 64.8 50.9 76.3

127 9.0 7.3 10.9



Abundance Ion 82.90 (82.60 to 83.60): VX0
Ion 84.90 (84.60 to 85.60): VX0
Ion 126.80 (126.50 to 127.50): V



#48

Methyl methacrylate

Concen: 46.654 ug/l

RT: 7.77 min Scan# 1086

Delta R.T. -0.01 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35


Tgt Ion: 41 Resp: 180796

Ion Ratio Lower Upper



41 100


69 84.1 70.2 105.4

39 50.8 42.7 64.1



Abundance Ion 41.00 (40.70 to 41.70): VX0
Ion 69.00 (68.70 to 69.70): VX0
Ion 39.00 (38.70 to 39.70): VX0









#49
1,4-Dioxane
Concen: 892.959 ug/l
RT: 7.76 min Scan# 1083
Delta R.T. -0.00 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tot Ion: 88 Resp: 79815
Ion Ratio Lower Upper
88 100
43 31.2 24.7 37.1
58 71.9 55.3 82.9



#50
Toluene-d8
Concen: 44.959 ug/l
RT: 8.71 min Scan# 1240
Delta R.T. -0.00 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35

Tgt Ion: 98 Resp: 460385
Ion Ratio Lower Upper
98 100
100 65.2 52.9 79.3





#51

4-Methyl-2-Pentanone

Concen: 236.946 ug/l

RT: 8.65 min Scan# 1230

Delta R.T. -0.00 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050EC

Tot Ion: 43 Resp: 1225868

Ion Ratio Lower Upper




43 100

58 39.4 33.1 49.7



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0



#52

Toluene

Concen: 47.506 ug/l

RT: 8.79 min Scan# 1252

Delta R.T. -0.00 min

Lab File: VX005396.D


Acq: 17 Oct 2018 17:35

Tgt Ion: 92 Resp: 340047

Ion Ratio Lower Upper



92 100


91 174.3 136.4 204.6



Abundance Ion 92.00 (91.70 to 92.70): VX0

Ion 91.00 (90.70 to 91.70): VX0









#53
 t-1,3-Dichloropropene
 Concen: 49.664 ug/l
 RT: 9.04 min Scan# 1294
 Delta R.T. -0.00 min
 Lab File: VX005396.D
 Acq: 17 Oct 2018 17:35




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tot Ion: 75 Resp: 210592
 Ion Ratio Lower Upper
 75 100
 77 31.1 24.6 37.0



#54
 cis-1,3-Dichloropropene
 Concen: 49.063 ug/l
 RT: 8.44 min Scan# 1195
 Delta R.T. -0.00 min
 Lab File: VX005396.D
 Acq: 17 Oct 2018 17:35

Tgt Ion: 75 Resp: 225422
 Ion Ratio Lower Upper
 75 100
 77 31.2 26.2 39.2
 39 43.7 36.4 54.6









#55
 1,1,2-Trichloroethane
 Concen: 46.991 ug/l
 RT: 9.22 min Scan# 1323
 Delta R.T. -0.00 min
 Lab File: VX005396.D
 Acq: 17 Oct 2018 17:35

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Tot Ion:	97	Resp:	142469
Ion	Ratio	Lower	Upper
97	100		
83	84.7	66.4	99.6
85	55.2	43.3	64.9
99	65.3	49.0	73.4




Abundance Ion 96.90 (96.60 to 97.60): VX0
 Ion 82.90 (82.60 to 83.60): VX0
 Ion 84.90 (84.60 to 85.60): VX0
 Ion 98.90 (98.60 to 99.60): VX0






#56
 Ethyl methacrylate
 Concen: 52.577 ug/l
 RT: 9.18 min Scan# 1317
 Delta R.T. -0.00 min
 Lab File: VX005396.D
 Acq: 17 Oct 2018 17:35

Tgt Ion:	69	Resp:	232380
Ion	Ratio	Lower	Upper
69	100		
41	67.2	51.0	76.4
39	31.0	26.1	39.1



Abundance Ion 69.00 (68.70 to 69.70): VX0
 Ion 41.00 (40.70 to 41.70): VX0
 Ion 39.00 (38.70 to 39.70): VX0





#57

1,3-Dichloropropane

Concen: 46.570 ug/l

RT: 9.37 min Scan# 1348

Delta R.T. -0.00 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050EC

Tot Ion: 76 Resp: 235662

Ion Ratio Lower Upper

76 100

78 31.9 25.5 38.3



Abundance Ion 75.90 (75.60 to 76.60): VX0

Ion 77.90 (77.60 to 78.60): VX0

9.37



150000

100000

50000

0

Time-->



#58

2-Chloroethyl Vinyl ether

Concen: 247.561 ug/l

RT: 8.32 min Scan# 1175

Delta R.T. -0.00 min

Lab File: VX005396.D


Acq: 17 Oct 2018 17:35

Tgt Ion: 63 Resp: 613634

Ion Ratio Lower Upper

63 100

106 27.4 23.9 35.9



Abundance Ion 62.90 (62.60 to 63.60): VX0

Ion 105.90 (105.60 to 106.60): V

8.32

400000


300000


200000

100000

0

Time-->









#59
2-Hexanone
Concen: 234.809 ug/l
RT: 9.50 min Scan# 1369
Delta R.T. -0.00 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tot Ion: 43 Resp: 977200
Ion Ratio Lower Upper
43 100
58 54.9 29.0 87.0



#60
Dibromochloromethane
Concen: 47.778 ug/l
RT: 9.59 min Scan# 1383
Delta R.T. -0.00 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35

Tgt Ion: 129 Resp: 149733
Ion Ratio Lower Upper
129 100
127 76.3 38.5 115.5





#61

1,2-Dibromoethane

Concen: 46.912 ug/l

RT: 9.67 min Scan# 1397

Delta R.T. -0.00 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050EC

Tot Ion:107 Resp: 149266

Ion Ratio Lower Upper




107 100

109 93.9 76.0 114.0



Abundance Ion 106.90 (106.60 to 107.60): V

Ion 108.80 (108.50 to 109.50): V



#62

4-Bromofluorobenzene

Concen: 45.734 ug/l

RT: 11.14 min Scan# 1638

Delta R.T. -0.00 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35


Tgt Ion: 95 Resp: 175930

Ion Ratio Lower Upper

95 100

174 94.3 0.0 185.2



176 92.8 0.0 178.6




Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.12 min Scan# 1470
Delta R.T. -0.00 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35

Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC

Tot Ion:117 Resp: 386246
Ion Ratio Lower Upper
117 100
82 52.4 47.8 71.6
119 32.1 29.3 43.9




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V






#64
Tetrachloroethene
Concen: 48.288 ug/l
RT: 9.34 min Scan# 1342
Delta R.T. -0.00 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35

Tgt Ion:164 Resp: 150150
Ion Ratio Lower Upper
164 100
166 125.6 102.8 154.2
129 89.9 69.4 104.0
131 85.7 67.9 101.9



Abundance Ion 163.80 (163.50 to 164.50): V
Ion 165.80 (165.50 to 166.50): V
Ion 128.80 (128.50 to 129.50): V
Ion 130.80 (130.50 to 131.50): V









#65
Chlorobenzene
Concen: 47.605 ug/l
RT: 10.14 min Scan# 1474
Delta R.T. -0.00 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tot Ion:112 Resp: 391122
Ion Ratio Lower Upper
112 100
114 31.6 27.8 41.8



#66
1,1,1,2-Tetrachloroethane
Concen: 49.041 ug/l
RT: 10.22 min Scan# 1487
Delta R.T. -0.01 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35

Tgt Ion:131 Resp: 143583
Ion Ratio Lower Upper
131 100
133 95.9 48.1 144.4
119 64.3 32.9 98.6





#67

Ethyl Benzene

Concen: 49.873 ug/l

RT: 10.25 min Scan# 1492

Delta R.T. -0.01 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tot Ion: 91 Resp: 666126

Ion Ratio Lower Upper

91 100

106 30.4 25.9 38.9



#68

m/p-Xylenes

Concen: 100.068 ug/l

RT: 10.36 min Scan# 1510

Delta R.T. -0.00 min

Lab File: VX005396.D




Acq: 17 Oct 2018 17:35


Tgt Ion: 106 Resp: 518759

Ion Ratio Lower Upper

106 100

91 202.4 157.1 235.7









#69
o-Xylene
Concen: 50.812 ug/l
RT: 10.70 min Scan# 1566
Delta R.T. -0.00 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tot Ion:106 Resp: 254047
Ion Ratio Lower Upper
106 100
91 213.8 105.1 315.1



#70
Styrene
Concen: 51.645 ug/l
RT: 10.71 min Scan# 1568
Delta R.T. -0.00 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35

Tgt Ion:104 Resp: 424701
Ion Ratio Lower Upper
104 100
78 49.1 37.4 56.0
103 55.6 43.8 65.6





#71

Bromoform

Concen: 49.187 ug/l

RT: 10.86 min Scan# 1592

Delta R.T. -0.00 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC


Tot Ion:173 Resp: 128692

Ion Ratio Lower Upper




173 100

175 48.8 24.1 72.3

254 0.2 0.2 0.2



Abundance Ion 172.70 (172.40 to 173.40): V
Ion 174.70 (174.40 to 175.40): V
Ion 254.40 (254.10 to 255.10): V



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.08 min Scan# 1792

Delta R.T. -0.00 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35


Tgt Ion:152 Resp: 218923

Ion Ratio Lower Upper



152 100


115 75.8 39.0 117.0

150 172.5 0.0 351.0



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V





#73

Isopropylbenzene

Concen: 52.807 ug/l

RT: 11.02 min Scan# 1618

Delta R.T. -0.00 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35

Instrument :

MSVOA_X

ClientSampleId :





VSTDCCC050EC

Tot Ion:105 Resp: 687251

Ion Ratio Lower Upper

105 100

120 26.9 13.5 40.4



#74

N-ethyl acetate

Concen: 52.890 ug/l

RT: 10.90 min Scan# 1598

Delta R.T. -0.01 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35

Tgt Ion: 43 Resp: 331992




Ion Ratio Lower Upper


43 100

70 41.4 37.4 56.0

55 24.1 21.8 32.8

61 24.0 20.6 30.8









#75
 1,1,2,2-Tetrachloroethane
 Concen: 47.685 ug/l
 RT: 11.27 min Scan# 1659
 Delta R.T. -0.00 min
 Lab File: VX005396.D
 Acq: 17 Oct 2018 17:35




Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC


Tot Ion: 83 Resp: 234104
 Ion Ratio Lower Upper
 83 100
 131 11.1 5.2 15.6
 85 65.0 32.3 96.8



#76
 1,2,3-Trichloropropane
 Concen: 65.044 ug/l
 RT: 11.30 min Scan# 1664
 Delta R.T. -0.00 min
 Lab File: VX005396.D
 Acq: 17 Oct 2018 17:35

Tgt Ion: 75 Resp: 276209
 Ion Ratio Lower Upper
 75 100
 77 33.2 20.2 60.6









#77
Bromobenzene
Concen: 49.927 ug/l
RT: 11.26 min Scan# 1657
Delta R.T. -0.00 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tot Ion:156 Resp: 185786
Ion Ratio Lower Upper
156 100
77 141.0 71.5 214.3
158 97.2 48.9 146.8



#78
n-propylbenzene
Concen: 52.277 ug/l
RT: 11.36 min Scan# 1674
Delta R.T. -0.00 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35

Tgt Ion: 91 Resp: 782902
Ion Ratio Lower Upper
91 100
120 24.0 11.9 35.9





#79

2-Chlorotoluene

Concen: 50.570 ug/l

RT: 11.42 min Scan# 1684

Delta R.T. -0.00 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050EC

Tot Ion: 91 Resp: 463826

Ion Ratio Lower Upper




91 100

126 35.5 17.9 53.7



Abundance Ion 91.00 (90.70 to 91.70): VX005284.D (-1680) (-)

Ion 125.90 (125.60 to 126.60): VX005284.D (-1680) (-)



#80

1,3,5-Trimethylbenzene

Concen: 52.725 ug/l

RT: 11.51 min Scan# 1698

Delta R.T. -0.00 min

Lab File: VX005396.D


Acq: 17 Oct 2018 17:35

Tgt Ion: 105 Resp: 587607

Ion Ratio Lower Upper



105 100


120 49.0 25.6 76.8



Abundance Ion 105.00 (104.70 to 105.70): VX005284.D (-1692) (-)

Ion 120.00 (119.70 to 120.70): VX005284.D (-1692) (-)





#81

trans-1,4-Dichloro-2-butene

Concen: 50.888 ug/l

RT: 11.07 min Scan# 1627

Delta R.T. -0.00 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC


Tot Ion: 75 Resp: 71119

Ion Ratio Lower Upper

75 100

53 94.1 80.1 120.1




89 47.7 37.2 55.8



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 53.00 (52.70 to 53.70): VX0

Ion 88.90 (88.60 to 89.60): VX0



#82

4-Chlorotoluene

Concen: 50.363 ug/l

RT: 11.51 min Scan# 1699

Delta R.T. -0.00 min

Lab File: VX005396.D


Acq: 17 Oct 2018 17:35

Tgt Ion: 91 Resp: 547574

Ion Ratio Lower Upper



91 100


126 31.1 15.5 46.5



Abundance Ion 91.00 (90.70 to 91.70): VX0

Ion 125.90 (125.60 to 126.60): V









#83
tert-Butylbenzene
Concen: 52.938 ug/l
RT: 11.77 min Scan# 1741
Delta R.T. -0.01 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tot Ion:119 Resp: 597446
Ion Ratio Lower Upper
119 100
91 55.5 27.0 81.0
134 22.7 11.7 35.1



#84
1,2,4-Trimethylbenzene
Concen: 52.184 ug/l
RT: 11.80 min Scan# 1747
Delta R.T. -0.01 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35

Tgt Ion:105 Resp: 598503
Ion Ratio Lower Upper
105 100
120 45.5 23.2 69.6





#85

sec-Butylbenzene

Concen: 53.800 ug/l

RT: 11.94 min Scan# 1770

Delta R.T. -0.00 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35

Instrument :

MSVOA_X

ClientSampleId :


VSTDCCC050EC

Tot Ion:105 Resp: 718542

Ion Ratio Lower Upper




105 100

134 20.6 10.5 31.5



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 134.00 (133.70 to 134.70): V



#86

p-Isopropyltoluene

Concen: 53.418 ug/l

RT: 12.07 min Scan# 1790

Delta R.T. -0.00 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35


Tgt Ion:119 Resp: 644836

Ion Ratio Lower Upper

119 100

134 26.8 13.4 40.2



91 22.3 10.9 32.7




Abundance Ion 119.00 (118.70 to 119.70): V

Ion 134.00 (133.70 to 134.70): V

Ion 91.00 (90.70 to 91.70): VX0





#87

1,3-Dichlorobenzene

Concen: 49.750 ug/l

RT: 12.02 min Scan# 1783

Delta R.T. -0.00 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC


Tot Ion:146 Resp: 345930

Ion Ratio Lower Upper

146 100

111 37.5 18.7 56.1




148 64.0 32.1 96.5



Abundance Ion 145.90 (145.60 to 146.60): V

Ion 110.90 (110.60 to 111.60): V

Ion 147.90 (147.60 to 148.60): V



#88

1,4-Dichlorobenzene

Concen: 49.211 ug/l

RT: 12.10 min Scan# 1795

Delta R.T. -0.00 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35


Tgt Ion:146 Resp: 350057

Ion Ratio Lower Upper

146 100

111 36.7 18.1 54.1



148 64.6 32.0 96.0




Abundance Ion 145.90 (145.60 to 146.60): V

Ion 110.90 (110.60 to 111.60): V

Ion 147.90 (147.60 to 148.60): V









#89
n-Butylbenzene
Concen: 52.659 ug/l
RT: 12.39 min Scan# 1843
Delta R.T. -0.00 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35




Instrument :
MSVOA_X
ClientSampleId :
VSTDCCC050EC


Tot Ion: 91 Resp: 577172
Ion Ratio Lower Upper
91 100
92 52.6 27.3 81.9
134 27.1 13.6 40.7



#90
Hexachloroethane
Concen: 49.519 ug/l
RT: 12.60 min Scan# 1877
Delta R.T. -0.00 min
Lab File: VX005396.D
Acq: 17 Oct 2018 17:35

Tgt Ion: 117 Resp: 103019
Ion Ratio Lower Upper
117 100
201 99.0 46.9 140.6





#91

1,2-Dichlorobenzene

Concen: 48.846 ug/l

RT: 12.39 min Scan# 1843

Delta R.T. -0.00 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC


Tot Ion:146 Resp: 346597

Ion Ratio Lower Upper

146 100

111 39.4 19.4 58.1




148 64.6 32.8 98.3



Abundance Ion 145.90 (145.60 to 146.60): V

Ion 110.90 (110.60 to 111.60): V

Ion 147.90 (147.60 to 148.60): V



#92

1,2-Dibromo-3-Chloropropane

Concen: 48.584 ug/l

RT: 13.00 min Scan# 1943

Delta R.T. -0.00 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35


Tgt Ion: 75 Resp: 58181

Ion Ratio Lower Upper

75 100

155 113.9 48.0 144.2



157 126.4 61.4 184.1




Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 154.80 (154.50 to 155.50): V

Ion 156.80 (156.50 to 157.50): V









#93
 1,2,4-Trichlorobenzene
 Concen: 53.731 ug/l
 RT: 13.65 min Scan# 2049
 Delta R.T. -0.00 min
 Lab File: VX005396.D
 Acq: 17 Oct 2018 17:35

Instrument :
 MSVOA_X
 ClientSampleId :
 VSTDCCC050EC

Tot Ion:180 Resp: 279721
 Ion Ratio Lower Upper
 180 100
 182 94.9 48.4 145.0
 145 27.7 14.3 42.9




Abundance Ion 179.80 (179.50 to 180.50): V
 Ion 181.80 (181.50 to 182.50): V
 Ion 144.80 (144.50 to 145.50): V






#94
 Hexachlorobutadiene
 Concen: 49.488 ug/l
 RT: 13.79 min Scan# 2072
 Delta R.T. -0.00 min
 Lab File: VX005396.D
 Acq: 17 Oct 2018 17:35

Tgt Ion:225 Resp: 136068
 Ion Ratio Lower Upper
 225 100
 223 62.7 30.1 90.5
 227 64.3 30.8 92.4



Abundance Ion 224.70 (224.40 to 225.40): V
 Ion 222.70 (222.40 to 223.40): V
 Ion 226.70 (226.40 to 227.40): V





#95

Naphthalene

Concen: 56.264 ug/l

RT: 13.83 min Scan# 2080

Delta R.T. -0.00 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35

Instrument :

MSVOA_X

ClientSampleId :

VSTDCCC050EC


Tot Ion:128 Resp: 849462

Ion Ratio Lower Upper

128 100

127 12.7 10.2 15.2




129 10.9 8.6 12.8



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 127.00 (126.70 to 127.70): V

Ion 129.00 (128.70 to 129.70): V



#96

1,2,3-Trichlorobenzene

Concen: 52.466 ug/l

RT: 14.02 min Scan# 2111

Delta R.T. -0.00 min

Lab File: VX005396.D

Acq: 17 Oct 2018 17:35


Tgt Ion:180 Resp: 279469

Ion Ratio Lower Upper

180 100

182 95.9 48.5 145.6

145 29.5 14.9 44.9



Abundance Ion 179.80 (179.50 to 180.50): V

Ion 181.80 (181.50 to 182.50): V

Ion 144.90 (144.60 to 145.60): V

