

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX101718\
 Data File : VX005390.D
 Acq On : 17 Oct 2018 14:55
 Operator : JC/MD
 Sample : J5572-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OUTFALL001-181016

Quant Time: Oct 17 17:37:43 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	284670	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	398587	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	356034	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	176861	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	161445	51.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.20%	
35) Dibromofluoromethane	5.50	113	126173	47.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.58%	
50) Toluene-d8	8.71	98	470920	51.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.92%	
62) 4-Bromofluorobenzene	11.14	95	156322	45.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.82%	
Target Compounds						
16) Acetone	2.44	43	17738	9.71	ug/l	88
27) cis-1,2-Dichloroethene	4.60	96	6288	2.09	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX101718\
Data File : VX005390.D
Acq On : 17 Oct 2018 14:55
Operator : JC/MD
Sample : J5572-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OUTFALL001-181016

Quant Time: Oct 17 17:37:43 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration

