


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101718\
Data File : VX005397.D
Acq On : 17 Oct 2018 18:28
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Title : SW846 8260
Last Update : Fri Oct 12 11:28:43 2018



AutoFind: Scans 1637, 1638, 1639; Background Corrected with Scan 1630

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.1	7877	PASS
75	95	30	60	46.9	21613	PASS
95	95	100	100	100.0	46069	PASS
96	95	5	9	6.9	3183	PASS
173	174	0.00	2	1.2	520	PASS
174	95	50	100	97.8	45058	PASS
175	174	5	9	7.4	3330	PASS
176	174	95	101	96.3	43413	PASS
177	176	5	9	6.8	2932	PASS