

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101719\
 Data File : VX013105.D
 Acq On : 17 Oct 2019 14:24
 Operator : JC/SP
 Sample : K5393-25
 Misc : 5.22g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 T-13-V1-(24)

Quant Time: Oct 18 10:08:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 16 07:26:22 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	161473	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	257403	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	213180	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	93828	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.05	65	95743	47.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.88%	
35) Dibromofluoromethane	5.49	113	72583	49.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.26%	
50) Toluene-d8	8.71	98	297037	51.55	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.10%	
62) 4-Bromofluorobenzene	11.14	95	102618	45.83	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.66%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101719\
Data File : VX013105.D
Acq On : 17 Oct 2019 14:24
Operator : JC/SP
Sample : K5393-25
Misc : 5.22g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
T-13-V1-(24)

Quant Time: Oct 18 10:08:21 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 16 07:26:22 2019
Response via : Initial Calibration

