

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101719\
 Data File : VX013111.D
 Acq On : 17 Oct 2019 16:44
 Operator : JC/SP
 Sample : K5406-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-3

Quant Time: Oct 18 09:11:27 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 16 07:26:22 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	177046	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	280932	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	234387	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	92372	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.05	65	106048	47.92	ug/l	0.00
Spiked Amount			Recovery	=	95.84%	
35) Dibromofluoromethane	5.49	113	81958	50.83	ug/l	0.00
Spiked Amount			Recovery	=	101.66%	
50) Toluene-d8	8.71	98	321672	51.15	ug/l	0.00
Spiked Amount			Recovery	=	102.30%	
62) 4-Bromofluorobenzene	11.14	95	102819	42.08	ug/l	0.00
Spiked Amount			Recovery	=	84.16%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101719\
Data File : VX013111.D
Acq On : 17 Oct 2019 16:44
Operator : JC/SP
Sample : K5406-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-3

Quant Time: Oct 18 09:11:27 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 16 07:26:22 2019
Response via : Initial Calibration

