

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101719\
 Data File : VX013115.D
 Acq On : 17 Oct 2019 18:18
 Operator : JC/SP
 Sample : K5414-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP-1-SRI4

Quant Time: Oct 18 08:45:21 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 16 07:26:22 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.65	168	161842	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	261335	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	214841	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	78490	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	98999	48.94	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	97.88%	
35) Dibromofluoromethane	5.49	113	77677	51.79	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	103.58%	
50) Toluene-d8	8.71	98	302247	51.67	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	103.34%	
62) 4-Bromofluorobenzene	11.13	95	94348	41.51	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	83.02%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.44	43	2610	2.448	ug/l	97
17) Carbon Disulfide	2.56	76	3033	0.650	ug/l	99
30) Chloroform	5.20	83	3285	1.020	ug/l	92
44) Trichloroethene	7.20	130	3850	2.046	ug/l	93
60) Dibromochloromethane	9.58	129	692	2.205	ug/l	73
64) Tetrachloroethene	9.33	164	9721	6.524	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101719\
Data File : VX013115.D
Acq On : 17 Oct 2019 18:18
Operator : JC/SP
Sample : K5414-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP-1-SRI4

Quant Time: Oct 18 08:45:21 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
Quant Title : SW846 8260
QLast Update : Wed Oct 16 07:26:22 2019
Response via : Initial Calibration

