


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX101719\
 Data File : VX013098.D
 Acq On : 17 Oct 2019 09:18
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X100819W.M
 Title : SW846 8260
 Last Update : Wed Oct 16 07:26:22 2019



AutoFind: Scans 1679, 1680, 1681; Background Corrected with Scan 1673

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.1	4693	PASS
75	95	30	60	51.1	13244	PASS
95	95	100	100	100.0	25917	PASS
96	95	5	9	7.7	1984	PASS
173	174	0.00	2	1.2	219	PASS
174	95	50	100	68.0	17626	PASS
175	174	5	9	8.4	1477	PASS
176	174	95	101	97.7	17229	PASS
177	176	5	9	7.1	1225	PASS