

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101722\
 Data File : VX032163.D
 Acq On : 17 Oct 2022 18:36
 Operator : JC/MD
 Sample : VX1017WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1017WBL01

Quant Time: Oct 18 06:12:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 04:18:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	127496	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	213937	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	156966	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86889	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	100481	49.483	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	98.960%
35) Dibromofluoromethane	5.385	113	72311	49.537	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.080%
50) Toluene-d8	8.647	98	263455	51.455	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.920%
62) 4-Bromofluorobenzene	11.079	95	71652	36.756	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	73.520%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101722\
Data File : VX032163.D
Acq On : 17 Oct 2022 18:36
Operator : JC/MD
Sample : VX1017WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1017WBL01

Quant Time: Oct 18 06:12:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101722W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 04:18:35 2022
Response via : Initial Calibration

