

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101722\
 Data File : VX032168.D
 Acq On : 17 Oct 2022 20:55
 Operator : JC/MD
 Sample : N5159-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-32

Quant Time: Oct 18 04:25:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 04:18:35 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	124648	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	217583	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	206228	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93434	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	103334	52.051	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.100%
35) Dibromofluoromethane	5.385	113	73584	49.564	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.120%
50) Toluene-d8	8.647	98	232825	44.711	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.420%
62) 4-Bromofluorobenzene	11.079	95	104698	52.808	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.620%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	21367	44.762	ug/l	98
20) Methylene Chloride	2.782	84	6431	4.382	ug/l #	87
43) Isopropyl Acetate	6.342	43	87760	27.019	ug/l	100
52) Toluene	8.720	92	5939	1.552	ug/l	100
95) Naphthalene	13.780	128	30059	5.099	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101722\
Data File : VX032168.D
Acq On : 17 Oct 2022 20:55
Operator : JC/MD
Sample : N5159-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-32

Quant Time: Oct 18 04:25:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101722W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 04:18:35 2022
Response via : Initial Calibration

