

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101722\
 Data File : VX032169.D
 Acq On : 17 Oct 2022 21:18
 Operator : JC/MD
 Sample : N5159-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-33

Quant Time: Oct 18 04:26:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 04:18:35 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	121455	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	210241	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	194534	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	87250	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	103385	53.445	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 106.900%	
35) Dibromofluoromethane	5.385	113	72892	50.813	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.620%	
50) Toluene-d8	8.647	98	262229	52.116	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 104.240%	
62) 4-Bromofluorobenzene	11.079	95	96720	50.487	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.980%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	15491	33.306	ug/l	93
20) Methylene Chloride	2.788	84	6329	4.426	ug/l	96
37) Ethyl Acetate	4.721	43	18714	8.986	ug/l	99
40) Benzene	6.044	78	2524	0.435	ug/l	94
43) Isopropyl Acetate	6.342	43	86580	27.587	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101722\
Data File : VX032169.D
Acq On : 17 Oct 2022 21:18
Operator : JC/MD
Sample : N5159-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-33

Quant Time: Oct 18 04:26:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101722W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 04:18:35 2022
Response via : Initial Calibration

