

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101722\
 Data File : VX032171.D
 Acq On : 17 Oct 2022 22:05
 Operator : JC/MD
 Sample : N5159-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-35

Quant Time: Oct 18 04:27:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 04:18:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	115274	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	204661	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	154495	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83380	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	103362	56.299	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 112.600%	
35) Dibromofluoromethane	5.385	113	73308	52.496	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.000%	
50) Toluene-d8	8.647	98	258080	52.690	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 105.380%	
62) 4-Bromofluorobenzene	11.079	95	74525	39.962	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 79.920%#	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	12535	28.395	ug/l	93
20) Methylene Chloride	2.788	84	6457	4.757	ug/l	95
37) Ethyl Acetate	4.721	43	19249	9.494	ug/l	99
43) Isopropyl Acetate	6.342	43	89498	29.294	ug/l	98
52) Toluene	8.720	92	4411	1.225	ug/l	99
95) Naphthalene	13.780	128	13928	2.648	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101722\
Data File : VX032171.D
Acq On : 17 Oct 2022 22:05
Operator : JC/MD
Sample : N5159-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-35

Quant Time: Oct 18 04:27:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101722W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 04:18:35 2022
Response via : Initial Calibration

