

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101722\
 Data File : VX032173.D
 Acq On : 17 Oct 2022 22:51
 Operator : JC/MD
 Sample : N5159-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-37

Quant Time: Oct 18 04:29:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 04:18:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	121490	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	175265	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	193061	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86131	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	68820	35.567	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	71.140%#
35) Dibromofluoromethane	5.385	113	69591	58.193	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	116.380%
50) Toluene-d8	8.647	98	250798	59.791	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	119.580%#
62) 4-Bromofluorobenzene	11.079	95	93128	58.313	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	116.620%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	13445	28.899	ug/l	95
20) Methylene Chloride	2.788	84	6695	4.680	ug/l	92
43) Isopropyl Acetate	6.342	43	70712	27.027	ug/l	99
52) Toluene	8.720	92	4044	1.312	ug/l	98
95) Naphthalene	13.780	128	10295	1.895	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101722\
Data File : VX032173.D
Acq On : 17 Oct 2022 22:51
Operator : JC/MD
Sample : N5159-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-37

Quant Time: Oct 18 04:29:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101722W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 04:18:35 2022
Response via : Initial Calibration

