

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101722\
 Data File : VX032178.D
 Acq On : 18 Oct 2022 00:47
 Operator : JC/MD
 Sample : N5159-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-42

Quant Time: Oct 18 04:34:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 04:18:35 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	122926	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	211228	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	197138	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	89902	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	100285	51.222	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 102.440%	
35) Dibromofluoromethane	5.385	113	71369	49.519	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.040%	
50) Toluene-d8	8.647	98	214845	42.500	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 85.000%#	
62) 4-Bromofluorobenzene	11.079	95	99891	51.899	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.800%	
Target Compounds						
						Qvalue
16) Acetone	2.386	43	14546	30.900	ug/l	99
20) Methylene Chloride	2.788	84	7602	5.252	ug/l	95
37) Ethyl Acetate	4.727	43	19905	9.513	ug/l	99
40) Benzene	6.037	78	5978	1.025	ug/l	96
43) Isopropyl Acetate	6.342	43	90309	28.641	ug/l	100
52) Toluene	8.720	92	4007	1.078	ug/l	87
95) Naphthalene	13.780	128	12222	2.155	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101722\
Data File : VX032178.D
Acq On : 18 Oct 2022 00:47
Operator : JC/MD
Sample : N5159-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-42

Quant Time: Oct 18 04:34:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101722W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 04:18:35 2022
Response via : Initial Calibration

