

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101722\
 Data File : VX032180.D
 Acq On : 18 Oct 2022 01:34
 Operator : JC/MD
 Sample : N5159-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-44

Quant Time: Oct 18 04:35:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 04:18:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	108426	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	147656	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	195662	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85558	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	73059	42.307	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	84.620%
35) Dibromofluoromethane	5.385	113	67148	66.649	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	133.300%#
50) Toluene-d8	8.647	98	218740	61.900	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	123.800%#
62) 4-Bromofluorobenzene	11.079	95	100773	74.899	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	149.800%#
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	14899	35.882	ug/l	96
20) Methylene Chloride	2.788	84	6806	5.331	ug/l	96
37) Ethyl Acetate	4.715	43	20488m	14.007	ug/l	
40) Benzene	6.044	78	8239	2.021	ug/l	99
43) Isopropyl Acetate	6.342	43	73525	33.357	ug/l	99
52) Toluene	8.720	92	6163	2.373	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101722\
Data File : VX032180.D
Acq On : 18 Oct 2022 01:34
Operator : JC/MD
Sample : N5159-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-44

Quant Time: Oct 18 04:35:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101722W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 04:18:35 2022
Response via : Initial Calibration

