

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101722\
 Data File : VX032181.D
 Acq On : 18 Oct 2022 01:57
 Operator : JC/MD
 Sample : N5159-14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-45

Quant Time: Oct 18 04:35:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 04:18:35 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	124916	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	218551	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	202313	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	91812	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	100668	50.599	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	101.200%
35) Dibromofluoromethane	5.385	113	71945	48.246	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.500%
50) Toluene-d8	8.647	98	214319	40.975	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	81.940%#
62) 4-Bromofluorobenzene	11.079	95	99030	49.728	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.460%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	20859	43.604	ug/l	93
20) Methylene Chloride	2.788	84	6833	4.646	ug/l	99
37) Ethyl Acetate	4.721	43	20069	9.270	ug/l	99
43) Isopropyl Acetate	6.342	43	91848	28.153	ug/l	99
52) Toluene	8.720	92	5875	1.528	ug/l	98
95) Naphthalene	13.780	128	19794	3.417	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101722\
Data File : VX032181.D
Acq On : 18 Oct 2022 01:57
Operator : JC/MD
Sample : N5159-14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-45

Quant Time: Oct 18 04:35:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101722W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 04:18:35 2022
Response via : Initial Calibration

