

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101722\
 Data File : VX032182.D
 Acq On : 18 Oct 2022 02:20
 Operator : JC/MD
 Sample : PB148361ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB148361ZHE#01

Quant Time: Oct 18 04:36:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 04:18:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	70792	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	115812	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	100983	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	59708	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	94882	84.153	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 168.300%#	
35) Dibromofluoromethane	5.385	113	66814	84.552	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 169.100%#	
50) Toluene-d8	8.647	98	188177	67.893	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 135.780%#	
62) 4-Bromofluorobenzene	11.079	95	66458	62.976	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 125.960%#	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	17035	62.837	ug/l	96
20) Methylene Chloride	2.788	84	7135	8.560	ug/l	96
37) Ethyl Acetate	4.714	43	18979	16.543	ug/l #	94
43) Isopropyl Acetate	6.342	43	90309	52.237	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101722\
Data File : VX032182.D
Acq On : 18 Oct 2022 02:20
Operator : JC/MD
Sample : PB148361ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB148361ZHE#01

Quant Time: Oct 18 04:36:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101722W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 04:18:35 2022
Response via : Initial Calibration

