

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101722\
 Data File : VX032183.D
 Acq On : 18 Oct 2022 02:44
 Operator : JC/MD
 Sample : PB148361ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB148361ZHE#02

Quant Time: Oct 18 04:36:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 04:18:35 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	70973	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	92710	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	128795	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79292	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	93865	83.038	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 166.080%#		
35) Dibromofluoromethane	5.385	113	67126	106.115	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 212.220%#		
50) Toluene-d8	8.647	98	244600	110.240	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 220.480%#		
62) 4-Bromofluorobenzene	11.079	95	69326	82.064	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 164.120%#		
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	12346	45.424	ug/l	100
20) Methylene Chloride	2.788	84	7043	8.428	ug/l	90
37) Ethyl Acetate	4.727	43	18323	19.951	ug/l	99
43) Isopropyl Acetate	6.342	43	85680	61.909	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101722\
Data File : VX032183.D
Acq On : 18 Oct 2022 02:44
Operator : JC/MD
Sample : PB148361ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB148361ZHE#02

Quant Time: Oct 18 04:36:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101722W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 04:18:35 2022
Response via : Initial Calibration

