

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101722\
 Data File : VX032184.D
 Acq On : 18 Oct 2022 03:07
 Operator : JC/MD
 Sample : PB148361ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB148361ZHE#03

Quant Time: Oct 18 04:37:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 04:18:35 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	65553	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	89993	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	124041	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	72962	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	92532	88.627	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 177.260%#	
35) Dibromofluoromethane	5.385	113	64833	105.584	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 211.160%#	
50) Toluene-d8	8.646	98	226741	105.277	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 210.560%#	
62) 4-Bromofluorobenzene	11.079	95	85414	104.161	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 208.320%#	
Target Compounds						
					Qvalue	
16) Acetone	2.385	43	16734	66.660	ug/l	99
20) Methylene Chloride	2.788	84	7675	9.944	ug/l	92
37) Ethyl Acetate	4.720	43	18369	20.605	ug/l	99
43) Isopropyl Acetate	6.342	43	88476	65.860	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101722\
Data File : VX032184.D
Acq On : 18 Oct 2022 03:07
Operator : JC/MD
Sample : PB148361ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB148361ZHE#03

Quant Time: Oct 18 04:37:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101722W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 04:18:35 2022
Response via : Initial Calibration

