

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101722\
 Data File : VX032185.D
 Acq On : 18 Oct 2022 03:30
 Operator : JC/MD
 Sample : PB148361ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB148361ZHE#04

Quant Time: Oct 18 14:37:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101722W.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 04:18:35 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	70271	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	116543	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	129084	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	74910	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	91562	81.810	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 163.620%#		
35) Dibromofluoromethane	5.385	113	65306	82.125	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 164.260%#		
50) Toluene-d8	8.647	98	239866	85.999	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 172.000%#		
62) 4-Bromofluorobenzene	11.079	95	90889	85.587	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 171.180%#		
Target Compounds						Qvalue
16) Acetone	2.385	43	10320	38.349	ug/l	93
20) Methylene Chloride	2.794	84	5350	6.466	ug/l	91
37) Ethyl Acetate	4.720	43	18271	15.826	ug/l	99
43) Isopropyl Acetate	6.342	43	87222	50.135	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101722\
Data File : VX032185.D
Acq On : 18 Oct 2022 03:30
Operator : JC/MD
Sample : PB148361ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB148361ZHE#04

Quant Time: Oct 18 14:37:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101722W.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 04:18:35 2022
Response via : Initial Calibration

