



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101722\
Data File : VX032148.D
Acq On : 17 Oct 2022 10:08
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101722W.M
Title : SW846 8260
Last Update : Tue Oct 18 04:18:35 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.8	9138	PASS
75	95	30	60	57.4	24101	PASS
95	95	100	100	100.0	41989	PASS
96	95	5	9	6.8	2858	PASS
173	174	0.00	2	0.6	191	PASS
174	95	50	100	73.7	30928	PASS
175	174	5	9	7.1	2204	PASS
176	174	95	101	96.1	29715	PASS
177	176	5	9	6.5	1922	PASS