

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101723\
 Data File : VX038319.D
 Acq On : 17 Oct 2023 11:04
 Operator : JC/MD
 Sample : VX1017WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1017WBL01

Quant Time: Oct 18 03:54:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	110982	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	199780	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	208759	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	109096	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	82067	45.974	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.940%
35) Dibromofluoromethane	5.385	113	64485	44.796	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.600%
50) Toluene-d8	8.647	98	258067	46.914	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.820%
62) 4-Bromofluorobenzene	11.079	95	97086	46.555	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101723\
Data File : VX038319.D
Acq On : 17 Oct 2023 11:04
Operator : JC/MD
Sample : VX1017WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1017WBL01

Quant Time: Oct 18 03:54:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

