

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101723\
 Data File : VX038328.D
 Acq On : 17 Oct 2023 14:45
 Operator : JC/MD
 Sample : PB156409ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB156409ZHE#01

Quant Time: Oct 18 03:59:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	130403	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	236199	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	258836	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	128108	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	93966	44.800	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.600%
35) Dibromofluoromethane	5.385	113	75033	44.086	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.180%
50) Toluene-d8	8.647	98	309376	47.570	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.140%
62) 4-Bromofluorobenzene	11.079	95	120749	48.974	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.940%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	21501	24.562	ug/l	94
20) Methylene Chloride	2.794	84	13396	8.136	ug/l #	82
37) Ethyl Acetate	4.721	43	3079	1.105	ug/l #	86
43) Isopropyl Acetate	6.342	43	107022	25.720	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101723\
Data File : VX038328.D
Acq On : 17 Oct 2023 14:45
Operator : JC/MD
Sample : PB156409ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB156409ZHE#01

Quant Time: Oct 18 03:59:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

