

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101723\
 Data File : VX038333.D
 Acq On : 17 Oct 2023 16:41
 Operator : JC/MD
 Sample : PB156409ZHE#06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB156409ZHE#06

Quant Time: Oct 18 04:01:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	121748	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	225762	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	253122	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	129874	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	89317	45.611	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.220%
35) Dibromofluoromethane	5.385	113	72442	44.532	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.060%
50) Toluene-d8	8.647	98	292578	47.067	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.140%
62) 4-Bromofluorobenzene	11.079	95	119835	50.850	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.700%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	20833	25.491	ug/l #	89
20) Methylene Chloride	2.788	84	15357	9.990	ug/l	91
37) Ethyl Acetate	4.727	43	3419m	1.283	ug/l	
43) Isopropyl Acetate	6.342	43	102639	25.807	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101723\
Data File : VX038333.D
Acq On : 17 Oct 2023 16:41
Operator : JC/MD
Sample : PB156409ZHE#06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB156409ZHE#06

Quant Time: Oct 18 04:01:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

