

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101723\
 Data File : VX038334.D
 Acq On : 17 Oct 2023 17:04
 Operator : JC/MD
 Sample : 04887-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SU-622-VOC-04

Quant Time: Oct 18 04:02:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	120708	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	218967	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	243379	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	121255	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	86756	44.685	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	89.360%
35) Dibromofluoromethane	5.385	113	69358	43.959	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.920%
50) Toluene-d8	8.647	98	285768	47.398	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.800%
62) 4-Bromofluorobenzene	11.079	95	115527	50.543	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.080%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	30635	37.808	ug/l	92
20) Methylene Chloride	2.788	84	10221	6.706	ug/l	89
43) Isopropyl Acetate	6.342	43	95439	24.741	ug/l	97
95) Naphthalene	13.774	128	95074	10.619	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101723\
Data File : VX038334.D
Acq On : 17 Oct 2023 17:04
Operator : JC/MD
Sample : 04887-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SU-622-VOC-04

Quant Time: Oct 18 04:02:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

