

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101723\
 Data File : VX038335.D
 Acq On : 17 Oct 2023 17:27
 Operator : JC/MD
 Sample : 04888-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TR-304-VOC-21

Quant Time: Oct 18 04:02:41 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	118160	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	216904	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	237956	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	118612	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	85893	45.194	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.380%
35) Dibromofluoromethane	5.391	113	68801	44.021	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.040%
50) Toluene-d8	8.647	98	281799	47.184	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.360%
62) 4-Bromofluorobenzene	11.079	95	110485	48.797	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.600%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	32312	40.737	ug/l	93
20) Methylene Chloride	2.794	84	13347	8.946	ug/l #	83
43) Isopropyl Acetate	6.342	43	70398	18.423	ug/l #	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101723\
Data File : VX038335.D
Acq On : 17 Oct 2023 17:27
Operator : JC/MD
Sample : 04888-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TR-304-VOC-21

Quant Time: Oct 18 04:02:41 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

