

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101723\
 Data File : VX038336.D
 Acq On : 17 Oct 2023 17:51
 Operator : JC/MD
 Sample : 04905-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LSS-B4-COMP-101323

Quant Time: Oct 18 04:03:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 05 15:27:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	116914	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	215044	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	236643	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	117217	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	85264	45.341	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.680%
35) Dibromofluoromethane	5.385	113	67622	43.640	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.280%
50) Toluene-d8	8.647	98	280296	47.338	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.680%
62) 4-Bromofluorobenzene	11.079	95	109681	48.861	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.720%
Target Compounds						
						Qvalue
16) Acetone	2.374	43	22369	28.502	ug/l	92
20) Methylene Chloride	2.788	84	12493	8.463	ug/l	92
37) Ethyl Acetate	4.715	43	2904	1.144	ug/l	# 90
43) Isopropyl Acetate	6.342	43	98002	25.869	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101723\
Data File : VX038336.D
Acq On : 17 Oct 2023 17:51
Operator : JC/MD
Sample : 04905-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LSS-B4-COMP-101323

Quant Time: Oct 18 04:03:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X100523W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 05 15:27:13 2023
Response via : Initial Calibration

