

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101725\
 Data File : VX048204.D
 Acq On : 17 Oct 2025 10:57
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Oct 18 00:10:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 18 00:06:31 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	80663	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	169165	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	177920	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	83522	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	70120	56.113	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	112.220%
35) Dibromofluoromethane	5.379	113	57791	49.224	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.440%
50) Toluene-d8	8.635	98	192615	45.540	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.080%#
62) 4-Bromofluorobenzene	11.061	95	87462	56.774	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	113.540%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.172	85	307	0.363	ug/l #	44
3) Chloromethane	1.307	50	909	0.886	ug/l	97
4) Vinyl Chloride	1.380	62	438	0.405	ug/l	94
5) Bromomethane	1.624	94	1203	1.674	ug/l	94
6) Chloroethane	1.703	64	183	0.266	ug/l #	45
7) Trichlorofluoromethane	1.898	101	492	0.290	ug/l	86
9) 1,1,2-Trichlorotrifluo...	2.337	101	431	0.451	ug/l #	80
10) Methyl Iodide	2.465	142	3320	2.900	ug/l #	83
11) Tert butyl alcohol	2.947	59	288	3.473	ug/l #	72
12) 1,1-Dichloroethene	2.331	96	405	0.433	ug/l #	45
13) Acrolein	2.245	56	588	3.374	ug/l #	76
14) Allyl chloride	2.666	41	628	0.374	ug/l #	67
15) Acrylonitrile	3.075	53	1274	3.207	ug/l	88
16) Acetone	2.380	43	974	2.285	ug/l	91
17) Carbon Disulfide	2.520	76	6131	2.132	ug/l	97
18) Methyl Acetate	2.703	43	539	0.621	ug/l #	81
20) Methylene Chloride	2.788	84	358	0.345	ug/l #	71
21) trans-1,2-Dichloroethene	3.099	96	785	0.812	ug/l	89
23) Vinyl Acetate	3.739	43	2426	0.985	ug/l #	73
24) 1,1-Dichloroethane	3.611	63	464	0.248	ug/l #	82
25) 2-Butanone	4.580	43	677	1.363	ug/l #	50
28) Bromochloromethane	4.897	49	293	0.340	ug/l #	83
29) Tetrahydrofuran	5.007	42	228	0.782	ug/l #	68
31) Cyclohexane	5.452	56	728	0.476	ug/l #	85
36) 1,1-Dichloropropene	5.696	75	502	0.304	ug/l #	64
38) Carbon Tetrachloride	5.659	117	488	0.257	ug/l #	61
39) Methylcyclohexane	7.360	83	1733	0.929	ug/l	84
40) Benzene	6.019	78	1192	0.243	ug/l #	82
42) 1,2-Dichloroethane	6.074	62	578	0.329	ug/l #	76
44) Trichloroethene	7.110	130	462	0.373	ug/l	92
46) Dibromomethane	7.574	93	270	0.302	ug/l #	71
49) 1,4-Dioxane	7.665	88	204	19.953	ug/l #	41
51) 4-Methyl-2-Pentanone	8.561	43	1450	1.174	ug/l	90
52) Toluene	8.708	92	844	0.286	ug/l	92
53) t-1,3-Dichloropropene	8.982	75	616	0.340	ug/l	91
54) cis-1,3-Dichloropropene	8.360	75	456	0.231	ug/l #	1

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101725\
 Data File : VX048204.D
 Acq On : 17 Oct 2025 10:57
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :

Quant Time: Oct 18 00:10:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 18 00:06:31 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


55) 1,1,2-Trichloroethane	9.147	97	296	0.269	ug/l #	58
57) 1,3-Dichloropropane	9.299	76	409	0.213	ug/l	98
58) 2-Chloroethyl Vinyl ether	8.250	63	543	14.884	ug/l #	55
59) 2-Hexanone	9.427	43	1019	1.180	ug/l	95
60) Dibromochloromethane	9.506	129	373	0.271	ug/l	92
61) 1,2-Dibromoethane	9.598	107	367	0.325	ug/l	91
64) Tetrachloroethene	9.256	164	1006	0.806	ug/l	87
65) Chlorobenzene	10.067	112	1588	0.408	ug/l #	85
67) Ethyl Benzene	10.183	91	2357	0.353	ug/l	94
68) m/p-Xylenes	10.287	106	1750	0.704	ug/l	88
69) o-Xylene	10.628	106	722	0.309	ug/l	98
70) Styrene	10.646	104	1576	0.391	ug/l	87
71) Bromoform	10.787	173	350	0.340	ug/l #	83
73) Isopropylbenzene	10.945	105	2582	0.444	ug/l	98
74) N-amyl acetate	10.835	43	574	0.265	ug/l #	84
75) 1,1,2,2-Tetrachloroethane	11.195	83	353	0.214	ug/l #	76
76) 1,2,3-Trichloropropane	11.061	75	46633	36.563	ug/l #	32
77) Bromobenzene	11.177	156	675	0.464	ug/l	84
78) n-propylbenzene	11.286	91	4917	0.698	ug/l	96
79) 2-Chlorotoluene	11.354	91	2373	0.558	ug/l	87
80) 1,3,5-Trimethylbenzene	11.433	105	3163	0.663	ug/l	94
81) trans-1,4-Dichloro-2-b...	11.061	75	46633	89.530	ug/l #	10
82) 4-Chlorotoluene	11.439	91	3625	0.725	ug/l	98
83) tert-Butylbenzene	11.695	119	3744	0.777	ug/l	98
84) 1,2,4-Trimethylbenzene	11.738	105	3212	0.676	ug/l	90
85) sec-Butylbenzene	11.872	105	8299	1.409	ug/l	99
86) p-Isopropyltoluene	11.994	119	7671	1.545	ug/l	96
87) 1,3-Dichlorobenzene	11.951	146	2735	1.006	ug/l	100
88) 1,4-Dichlorobenzene	11.951	146	2735	0.971	ug/l	99
89) n-Butylbenzene	12.317	91	12918	2.794	ug/l	97
90) Hexachloroethane	12.518	117	1759	1.679	ug/l	91
91) 1,2-Dichlorobenzene	12.317	146	2156	0.839	ug/l	95
92) 1,2-Dibromo-3-Chloropr...	12.926	75	214	0.712	ug/l	81
93) 1,2,4-Trichlorobenzene	13.573	180	5218	3.441	ug/l	96
94) Hexachlorobutadiene	13.707	225	8817	15.077	ug/l	97
95) Naphthalene	13.756	128	11183	2.798	ug/l	98
96) 1,2,3-Trichlorobenzene	13.945	180	5503	3.933	ug/l	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101725\
 Data File : VX048204.D
 Acq On : 17 Oct 2025 10:57
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :


Quant Time: Oct 18 00:10:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 18 00:06:31 2025
 Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.544 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:168 Resp: 80663
Ion Ratio Lower Upper
168 100
99 60.6 46.4 69.6



#2
Dichlorodifluoromethane
Concen: 0.363 ug/l
RT: 1.172 min Scan# 14
Delta R.T. -0.000 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57

Tgt Ion: 85 Resp: 307
Ion Ratio Lower Upper
85 100
87 0.0 15.3 45.9#









#3
Chloromethane
Concen: 0.886 ug/l
RT: 1.307 min Scan# 30
Delta R.T. 0.006 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 50 Resp: 909
Ion Ratio Lower Upper
50 100
52 30.8 25.9 38.9



#4
Vinyl Chloride
Concen: 0.405 ug/l
RT: 1.380 min Scan# 48
Delta R.T. -0.000 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57

Tgt Ion: 62 Resp: 438
Ion Ratio Lower Upper
62 100
64 28.1 25.1 37.7









#5
Bromomethane
Concen: 1.674 ug/l
RT: 1.624 min Scan# 88
Delta R.T. 0.006 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 94 Resp: 1203
Ion Ratio Lower Upper
94 100
96 100.4 76.0 114.0



#6
Chloroethane
Concen: 0.266 ug/l
RT: 1.703 min Scan# 101
Delta R.T. 0.006 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57


Tgt Ion: 64 Resp: 183
Ion Ratio Lower Upper
64 100
66 0.0 23.9 35.9#








#7
 Trichlorofluoromethane
 Concen: 0.290 ug/l
 RT: 1.898 min Scan# 11
 Delta R.T. -0.000 min
 Lab File: VX048204.D
 Acq: 17 Oct 2025 10:57

Instrument :
 MSVOA_X
 ClientSampleId :







Tgt Ion:101 Resp: 492
 Ion Ratio Lower Upper
 101 100
 103 75.4 51.6 77.4



#9
 1,1,2-Trichlorotrifluoroethane
 Concen: 0.451 ug/l
 RT: 2.337 min Scan# 205
 Delta R.T. -0.000 min
 Lab File: VX048204.D
 Acq: 17 Oct 2025 10:57


Tgt Ion:101 Resp: 431
 Ion Ratio Lower Upper
 101 100
 85 41.1 35.6 53.4
 151 97.2 58.0 87.0#








#10
Methyl Iodide
Concen: 2.900 ug/l
RT: 2.465 min Scan# 21
Delta R.T. 0.006 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:142 Resp: 3320
Ion Ratio Lower Upper
142 100
127 37.1 38.5 57.7#
141 6.1 11.6 17.4#



#11
Tert butyl alcohol
Concen: 3.473 ug/l
RT: 2.947 min Scan# 305
Delta R.T. 0.006 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57

Tgt Ion: 59 Resp: 288
Ion Ratio Lower Upper
59 100
57 0.0 8.2 12.4#





#12

1,1-Dichloroethene

Concen: 0.433 ug/l

RT: 2.331 min Scan# 20

Delta R.T. 0.006 min

Lab File: VX048204.D

Acq: 17 Oct 2025 10:57

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion: 96 Resp: 405

Ion Ratio Lower Upper

96 100

61 85.7 139.0 208.6#

98 40.9 52.4 78.6#



#13

Acrolein

Concen: 3.374 ug/l

RT: 2.245 min Scan# 190

Delta R.T. 0.012 min

Lab File: VX048204.D




Acq: 17 Oct 2025 10:57


Tgt Ion: 56 Resp: 588

Ion Ratio Lower Upper

56 100


55 50.0 56.1 84.1#








#14
 Allyl chloride
 Concen: 0.374 ug/l
 RT: 2.666 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: VX048204.D
 Acq: 17 Oct 2025 10:57


Instrument :
 MSVOA_X
 ClientSampleId :





Tgt Ion: 41 Resp: 628
 Ion Ratio Lower Upper
 41 100
 39 86.5 56.0 84.0#
 76 0.0 27.5 41.3#




#15
 Acrylonitrile
 Concen: 3.207 ug/l
 RT: 3.075 min Scan# 326
 Delta R.T. 0.018 min
 Lab File: VX048204.D
 Acq: 17 Oct 2025 10:57



Tgt Ion: 53 Resp: 1274
 Ion Ratio Lower Upper
 53 100
 52 68.5 66.1 99.1
 51 39.2 29.0 43.4









#16
Acetone
Concen: 2.285 ug/l
RT: 2.380 min Scan# 211
Delta R.T. 0.006 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 974
Ion Ratio Lower Upper
43 100
58 34.0 23.4 35.2



#17
Carbon Disulfide
Concen: 2.132 ug/l
RT: 2.520 min Scan# 235
Delta R.T. 0.006 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57

Tgt Ion: 76 Resp: 6131
Ion Ratio Lower Upper
76 100
78 10.0 7.1 10.7









#18
Methyl Acetate
Concen: 0.621 ug/l
RT: 2.703 min Scan# 20
Delta R.T. 0.006 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 539
Ion Ratio Lower Upper
43 100
74 13.0 17.5 26.3#



#20
Methylene Chloride
Concen: 0.345 ug/l
RT: 2.788 min Scan# 279
Delta R.T. -0.000 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57

Tgt Ion: 84 Resp: 358
Ion Ratio Lower Upper
84 100
49 82.2 104.1 156.1#
51 26.8 31.3 46.9#
86 57.6 51.3 76.9









#21
trans-1,2-Dichloroethene
Concen: 0.812 ug/l
RT: 3.099 min Scan# 31
Delta R.T. 0.006 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 96 Resp: 785
Ion Ratio Lower Upper
96 100
61 121.8 111.9 167.9
98 60.9 51.2 76.8



#23
Vinyl Acetate
Concen: 0.985 ug/l
RT: 3.739 min Scan# 435
Delta R.T. 0.024 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57


Tgt Ion: 43 Resp: 2426
Ion Ratio Lower Upper
43 100
86 0.0 7.9 11.9#








#24
 1,1-Dichloroethane
 Concen: 0.248 ug/l
 RT: 3.611 min Scan# 41
 Delta R.T. 0.006 min
 Lab File: VX048204.D
 Acq: 17 Oct 2025 10:57

Instrument :
 MSVOA_X
 ClientSampleId :







Tgt Ion: 63 Resp: 464
 Ion Ratio Lower Upper
 63 100
 98 0.0 3.5 10.6#
 100 0.0 2.1 6.3#



#25
 2-Butanone
 Concen: 1.363 ug/l
 RT: 4.580 min Scan# 573
 Delta R.T. 0.043 min
 Lab File: VX048204.D
 Acq: 17 Oct 2025 10:57


Tgt Ion: 43 Resp: 677
 Ion Ratio Lower Upper
 43 100
 72 0.0 19.9 29.9#








#28
Bromochloromethane
Concen: 0.340 ug/l
RT: 4.897 min Scan# 61
Delta R.T. 0.012 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion: 49 Resp: 293
Ion Ratio Lower Upper
49 100
129 0.0 0.0 4.4
130 59.4 59.3 88.9




#29
Tetrahydrofuran
Concen: 0.782 ug/l
RT: 5.007 min Scan# 643
Delta R.T. 0.018 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57




Tgt Ion: 42 Resp: 228
Ion Ratio Lower Upper
42 100
72 19.3 35.3 52.9#
71 25.0 32.2 48.4#








#31
Cyclohexane
Concen: 0.476 ug/l
RT: 5.452 min Scan# 716
Delta R.T. -0.006 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 56 Resp: 728
Ion Ratio Lower Upper
56 100
69 34.9 23.0 34.6#
84 66.8 64.6 97.0



#33
1,2-Dichloroethane-d4
Concen: 56.113 ug/l
RT: 5.940 min Scan# 796
Delta R.T. -0.000 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57


Tgt Ion: 65 Resp: 70120
Ion Ratio Lower Upper
65 100
67 50.5 0.0 105.0








#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.745 min Scan# 91
Delta R.T. -0.000 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:114 Resp: 169165
Ion Ratio Lower Upper
114 100
63 22.4 0.0 43.2
88 16.7 0.0 33.6




#35
Dibromofluoromethane
Concen: 49.224 ug/l
RT: 5.379 min Scan# 704
Delta R.T. 0.006 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57



Tgt Ion:113 Resp: 57791
Ion Ratio Lower Upper
113 100
111 103.8 82.2 123.4
192 17.8 15.1 22.7





#36

1,1-Dichloropropene

Concen: 0.304 ug/l

RT: 5.696 min Scan# 71

Delta R.T. 0.012 min

Lab File: VX048204.D





Acq: 17 Oct 2025 10:57

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion:	75	Resp:	502
Ion	Ratio	Lower	Upper
75	100		
110	4.2	17.7	53.1#
77	22.3	24.3	36.5#



#38

Carbon Tetrachloride

Concen: 0.257 ug/l




RT: 5.659 min Scan# 750


Delta R.T. -0.006 min

Lab File: VX048204.D

Acq: 17 Oct 2025 10:57


Tgt Ion:	117	Resp:	488
Ion	Ratio	Lower	Upper
117	100		
119	63.5	77.2	115.8#
121	0.0	24.8	37.2#








#39
Methylcyclohexane
Concen: 0.929 ug/l
RT: 7.360 min Scan# 11
Delta R.T. -0.006 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion:	83	Resp:	1733
Ion	Ratio	Lower	Upper
83	100		
55	98.8	66.2	99.2
98	40.6	38.9	58.3



#40
Benzene
Concen: 0.243 ug/l
RT: 6.019 min Scan# 809
Delta R.T. -0.006 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57

Tgt Ion:	78	Resp:	1192
Ion	Ratio	Lower	Upper
78	100		
77	15.4	19.3	28.9#





#42

1,2-Dichloroethane

Concen: 0.329 ug/l

RT: 6.074 min Scan# 817

Delta R.T. 0.006 min

Lab File: VX048204.D

Acq: 17 Oct 2025 10:57

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 62 Resp: 578

Ion Ratio Lower Upper

62 100

98 0.0 0.0 17.2



#44

Trichloroethene

Concen: 0.373 ug/l

RT: 7.110 min Scan# 988

Delta R.T. -0.000 min

Lab File: VX048204.D




Acq: 17 Oct 2025 10:57


Tgt Ion:130 Resp: 462

Ion Ratio Lower Upper

130 100


95 108.4 0.0 200.6








#46
Dibromomethane
Concen: 0.302 ug/l
RT: 7.574 min Scan# 1064
Delta R.T. 0.006 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57

Instrument :
MSVOA_X
ClientSampleId :







Tgt Ion: 93 Resp: 270
Ion Ratio Lower Upper
93 100
95 39.3 65.9 98.9#
174 101.1 71.7 107.5



#49
1,4-Dioxane
Concen: 19.953 ug/l
RT: 7.665 min Scan# 1079
Delta R.T. 0.024 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57

Tgt Ion: 88 Resp: 204
Ion Ratio Lower Upper
88 100
43 0.0 35.0 52.4#
58 30.4 61.9 92.9#





#50

Toluene-d8

Concen: 45.540 ug/l

RT: 8.635 min Scan# 11

Delta R.T. -0.000 min

Lab File: VX048204.D

Acq: 17 Oct 2025 10:57

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 98 Resp: 192615

Ion Ratio Lower Upper

98 100

100 67.4 53.0 79.4



#51

4-Methyl-2-Pentanone

Concen: 1.174 ug/l

RT: 8.561 min Scan# 1226

Delta R.T. 0.006 min

Lab File: VX048204.D




Acq: 17 Oct 2025 10:57


Tgt Ion: 43 Resp: 1450

Ion Ratio Lower Upper

43 100

58 31.2 29.5 44.3









#52
Toluene
Concen: 0.286 ug/l
RT: 8.708 min Scan# 1249
Delta R.T. 0.006 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 92 Resp: 844
Ion Ratio Lower Upper
92 100
91 180.5 136.1 204.1



#53
t-1,3-Dichloropropene
Concen: 0.340 ug/l
RT: 8.982 min Scan# 1295
Delta R.T. 0.018 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57

Tgt Ion: 75 Resp: 616
Ion Ratio Lower Upper
75 100
77 26.2 25.0 37.6









#54
 cis-1,3-Dichloropropene
 Concen: 0.231 ug/l
 RT: 8.360 min Scan# 1193
 Delta R.T. 0.012 min
 Lab File: VX048204.D
 Acq: 17 Oct 2025 10:57




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 75 Resp: 456
 Ion Ratio Lower Upper
 75 100
 77 47.7 24.9 37.3#
 39 634.0 42.6 64.0#



#55
 1,1,2-Trichloroethane
 Concen: 0.269 ug/l
 RT: 9.147 min Scan# 1322
 Delta R.T. 0.012 min
 Lab File: VX048204.D
 Acq: 17 Oct 2025 10:57

Tgt Ion: 97 Resp: 296
 Ion Ratio Lower Upper
 97 100
 83 72.6 70.6 105.8
 85 0.0 45.0 67.6#
 99 28.4 50.9 76.3#





#57

1,3-Dichloropropane

Concen: 0.213 ug/l

RT: 9.299 min Scan# 1134

Delta R.T. 0.006 min

Lab File: VX048204.D

Acq: 17 Oct 2025 10:57

Instrument :

MSVOA_X





ClientSampleId :

Tgt Ion: 76 Resp: 409

Ion Ratio Lower Upper

76 100

78 30.3 25.4 38.0



#58

2-Chloroethyl Vinyl ether

Concen: 14.884 ug/l

RT: 8.250 min Scan# 1175

Delta R.T. 0.024 min

Lab File: VX048204.D




Acq: 17 Oct 2025 10:57


Tgt Ion: 63 Resp: 543

Ion Ratio Lower Upper

63 100

106 4.1 22.2 33.2#









#59
2-Hexanone
Concen: 1.180 ug/l
RT: 9.427 min Scan# 1368
Delta R.T. 0.012 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 43 Resp: 1019
Ion Ratio Lower Upper
43 100
58 56.2 26.3 78.8



#60
Dibromochloromethane
Concen: 0.271 ug/l
RT: 9.506 min Scan# 1381
Delta R.T. -0.000 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57

Tgt Ion: 129 Resp: 373
Ion Ratio Lower Upper
129 100
127 70.0 38.6 115.7









#61
1,2-Dibromoethane
Concen: 0.325 ug/l
RT: 9.598 min Scan# 1395
Delta R.T. 0.006 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:107 Resp: 367
Ion Ratio Lower Upper
107 100
109 87.7 77.0 115.4



#62
4-Bromofluorobenzene
Concen: 56.774 ug/l
RT: 11.061 min Scan# 1636
Delta R.T. -0.000 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57


Tgt Ion: 95 Resp: 87462
Ion Ratio Lower Upper
95 100
174 73.0 0.0 147.8
176 69.8 0.0 142.8








#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.037 min Scan# 1468
Delta R.T. -0.000 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:117 Resp: 177920
Ion Ratio Lower Upper
117 100
82 56.6 46.5 69.7
119 31.2 25.9 38.9




#64
Tetrachloroethene
Concen: 0.806 ug/l
RT: 9.256 min Scan# 1340
Delta R.T. -0.000 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57



Tgt Ion:164 Resp: 1006
Ion Ratio Lower Upper
164 100
166 104.9 104.5 156.7
129 105.4 81.8 122.6
131 105.7 77.0 115.6









#65
Chlorobenzene
Concen: 0.408 ug/l
RT: 10.067 min Scan# 1472
Delta R.T. 0.006 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:112 Resp: 1588
Ion Ratio Lower Upper
112 100
114 24.4 26.2 39.2#



#67
Ethyl Benzene
Concen: 0.353 ug/l
RT: 10.183 min Scan# 1492
Delta R.T. 0.006 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57

Tgt Ion: 91 Resp: 2357
Ion Ratio Lower Upper
91 100
106 32.7 23.8 35.6









#68
m/p-Xylenes
Concen: 0.704 ug/l
RT: 10.287 min Scan# 1509
Delta R.T. -0.000 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:106 Resp: 1750
Ion Ratio Lower Upper
106 100
91 228.1 167.9 251.9



#69
o-Xylene
Concen: 0.309 ug/l
RT: 10.628 min Scan# 1565
Delta R.T. 0.006 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57

Tgt Ion:106 Resp: 722
Ion Ratio Lower Upper
106 100
91 226.2 111.3 333.9





#70

Styrene

Concen: 0.391 ug/l

RT: 10.646 min Scan# 1568

Delta R.T. 0.006 min

Lab File: VX048204.D





Acq: 17 Oct 2025 10:57

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion:104	Resp:	1576
Ion Ratio	Lower	Upper
104	100	
78	43.4	42.2 63.4
103	44.0	43.0 64.4



#71

Bromoform

Concen: 0.340 ug/l




RT: 10.787 min Scan# 1591


Delta R.T. 0.006 min

Lab File: VX048204.D

Acq: 17 Oct 2025 10:57


Tgt Ion:173	Resp:	350
Ion Ratio	Lower	Upper
173	100	
175	36.9	24.0 72.0
254	0.0	0.0 0.0








#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.006 min Scan# 1791
Delta R.T. -0.000 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:152 Resp: 83522
Ion Ratio Lower Upper
152 100
115 63.1 43.1 129.4
150 155.5 0.0 353.0




#73
Isopropylbenzene
Concen: 0.444 ug/l
RT: 10.945 min Scan# 1617
Delta R.T. -0.000 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57



Tgt Ion:105 Resp: 2582
Ion Ratio Lower Upper
105 100
120 24.9 13.0 38.9





#74

N-amy1 acetate

Concen: 0.265 ug/l

RT: 10.835 min Scan# 1599

Delta R.T. 0.012 min

Lab File: VX048204.D

Acq: 17 Oct 2025 10:57

Instrument :

MSVOA_X

ClientSampleId :

Tgt Ion: 43 Resp: 574





Ion Ratio Lower Upper

43 100

70 43.9 32.2 48.2

55 25.3 19.1 28.7

61 0.0 18.6 28.0#



#75

1,1,2,2-Tetrachloroethane

Concen: 0.214 ug/l

RT: 11.195 min Scan# 1658

Delta R.T. -0.000 min

Lab File: VX048204.D

Acq: 17 Oct 2025 10:57




Tgt Ion: 83 Resp: 353


Ion Ratio Lower Upper

83 100

131 0.0 4.8 14.4#

85 45.3 31.9 95.7









#76
1,2,3-Trichloropropane
Concen: 36.563 ug/l
RT: 11.061 min Scan# 11
Delta R.T. -0.159 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 75 Resp: 46633
Ion Ratio Lower Upper
75 100
77 0.5 22.4 67.2#



#77
Bromobenzene
Concen: 0.464 ug/l
RT: 11.177 min Scan# 1655
Delta R.T. -0.000 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57

Tgt Ion:156 Resp: 675
Ion Ratio Lower Upper
156 100
77 198.4 87.0 261.0
158 83.3 47.8 143.4









#78
n-propylbenzene
Concen: 0.698 ug/l
RT: 11.286 min Scan# 1673
Delta R.T. -0.000 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 4917
Ion Ratio Lower Upper
91 100
120 24.2 11.1 33.1



#79
2-Chlorotoluene
Concen: 0.558 ug/l
RT: 11.354 min Scan# 1684
Delta R.T. 0.006 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57

Tgt Ion: 91 Resp: 2373
Ion Ratio Lower Upper
91 100
126 25.7 16.4 49.4









#80
1,3,5-Trimethylbenzene
Concen: 0.663 ug/l
RT: 11.433 min Scan# 1
Delta R.T. -0.000 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:105 Resp: 3163
Ion Ratio Lower Upper
105 100
120 51.8 24.0 72.0



#81
trans-1,4-Dichloro-2-butene
Concen: 89.530 ug/l
RT: 11.061 min Scan# 1636
Delta R.T. 0.061 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57

Tgt Ion: 75 Resp: 46633
Ion Ratio Lower Upper
75 100
53 0.0 79.5 119.3#
89 0.0 36.8 55.2#









#82
4-Chlorotoluene
Concen: 0.725 ug/l
RT: 11.439 min Scan# 1
Delta R.T. 0.006 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion: 91 Resp: 3625
Ion Ratio Lower Upper
91 100
126 27.3 14.1 42.4



#83
tert-Butylbenzene
Concen: 0.777 ug/l
RT: 11.695 min Scan# 1740
Delta R.T. -0.000 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57

Tgt Ion: 119 Resp: 3744
Ion Ratio Lower Upper
119 100
91 62.5 30.3 91.0
134 22.9 11.4 34.2









#84
1,2,4-Trimethylbenzene
Concen: 0.676 ug/l
RT: 11.738 min Scan# 1174
Delta R.T. 0.006 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:105 Resp: 3212
Ion Ratio Lower Upper
105 100
120 38.1 22.1 66.5



#85
sec-Butylbenzene
Concen: 1.409 ug/l
RT: 11.872 min Scan# 1769
Delta R.T. -0.000 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57

Tgt Ion:105 Resp: 8299
Ion Ratio Lower Upper
105 100
134 20.6 10.1 30.1









#86
p-Isopropyltoluene
Concen: 1.545 ug/l
RT: 11.994 min Scan# 1789
Delta R.T. 0.006 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:119 Resp: 7671
Ion Ratio Lower Upper
119 100
134 26.4 13.0 39.0
91 28.1 12.1 36.3



#87
1,3-Dichlorobenzene
Concen: 1.006 ug/l
RT: 11.951 min Scan# 1782
Delta R.T. -0.000 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57


Tgt Ion:146 Resp: 2735
Ion Ratio Lower Upper
146 100
111 40.9 20.4 61.2
148 64.3 31.9 95.9








#88
1,4-Dichlorobenzene
Concen: 0.971 ug/l
RT: 11.951 min Scan# 1794
Delta R.T. -0.073 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57


Instrument :
MSVOA_X
ClientSampleId :





Tgt Ion:146 Resp: 2735
Ion Ratio Lower Upper
146 100
111 40.9 20.1 60.2
148 64.3 31.8 95.3




#89
n-Butylbenzene
Concen: 2.794 ug/l
RT: 12.317 min Scan# 1842
Delta R.T. 0.006 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57



Tgt Ion: 91 Resp: 12918
Ion Ratio Lower Upper
91 100
92 51.4 26.9 80.6
134 23.9 12.7 38.1









#90
Hexachloroethane
Concen: 1.679 ug/l
RT: 12.518 min Scan# 1875
Delta R.T. -0.000 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57




Instrument :
MSVOA_X
ClientSampleId :


Tgt Ion:117 Resp: 1759
Ion Ratio Lower Upper
117 100
201 60.8 34.1 102.2



#91
1,2-Dichlorobenzene
Concen: 0.839 ug/l
RT: 12.317 min Scan# 1842
Delta R.T. -0.000 min
Lab File: VX048204.D
Acq: 17 Oct 2025 10:57

Tgt Ion:146 Resp: 2156
Ion Ratio Lower Upper
146 100
111 41.1 21.9 65.5
148 58.7 31.9 95.5









#92
 1,2-Dibromo-3-Chloropropane
 Concen: 0.712 ug/l
 RT: 12.926 min Scan# 1942
 Delta R.T. 0.006 min
 Lab File: VX048204.D
 Acq: 17 Oct 2025 10:57




Instrument :
 MSVOA_X
 ClientSampleId :


Tgt Ion: 75 Resp: 214
 Ion Ratio Lower Upper
 75 100
 155 96.3 42.8 128.4
 157 82.2 54.2 162.6



#93
 1,2,4-Trichlorobenzene
 Concen: 3.441 ug/l
 RT: 13.573 min Scan# 2048
 Delta R.T. 0.006 min
 Lab File: VX048204.D
 Acq: 17 Oct 2025 10:57

Tgt Ion: 180 Resp: 5218
 Ion Ratio Lower Upper
 180 100
 182 98.5 47.3 142.0
 145 35.9 16.3 48.9





#94

Hexachlorobutadiene

Concen: 15.077 ug/l

RT: 13.707 min Scan# 2069

Delta R.T. 0.006 min

Lab File: VX048204.D

Acq: 17 Oct 2025 10:57

Instrument :

MSVOA_X

ClientSampleId :





Tgt Ion:225 Resp: 8817

Ion Ratio Lower Upper

225 100

223 68.3 32.6 97.7

227 66.2 32.5 97.4



#95

Naphthalene

Concen: 2.798 ug/l

RT: 13.756 min Scan# 2078

Delta R.T. -0.000 min

Lab File: VX048204.D

Acq: 17 Oct 2025 10:57




Tgt Ion:128 Resp: 11183


Ion Ratio Lower Upper

128 100

127 12.1 10.2 15.4


129 10.1 8.6 12.8





#96
 1,2,3-Trichlorobenzene
 Concen: 3.933 ug/l
 RT: 13.945 min Scan# 2109
 Delta R.T. 0.006 min
 Lab File: VX048204.D
 Acq: 17 Oct 2025 10:57

Instrument :
 MSVOA_X
 ClientSampleId :



Tgt Ion:180 Resp: 5503
 Ion Ratio Lower Upper
 180 100
 182 95.7 46.6 139.7
 145 34.2 16.9 50.7

