

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101725\
 Data File : VX048206.D
 Acq On : 17 Oct 2025 12:54
 Operator : JC/MD
 Sample : VX1017MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1017MBL01

Quant Time: Oct 18 00:16:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 18 00:06:31 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.537	168	81483	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	166025	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	177931	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	86757	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.940	65	65400	51.809	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	103.620%
35) Dibromofluoromethane	5.373	113	57562	49.956	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.920%
50) Toluene-d8	8.634	98	212870	51.280	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.560%
62) 4-Bromofluorobenzene	11.061	95	89164	58.973	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	117.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101725\
Data File : VX048206.D
Acq On : 17 Oct 2025 12:54
Operator : JC/MD
Sample : VX1017MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1017MBL01

Quant Time: Oct 18 00:16:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101725W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 18 00:06:31 2025
Response via : Initial Calibration

