

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101725\
 Data File : VX048208.D
 Acq On : 17 Oct 2025 13:43
 Operator : JC/MD
 Sample : Q3311-12DL 5X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-214DDL

Quant Time: Oct 18 00:17:04 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 18 00:06:31 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.538	168	83059	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	168187	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	177290	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	85687	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.934	65	74775	58.112	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 116.220%	
35) Dibromofluoromethane	5.367	113	62512	53.554	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.100%	
50) Toluene-d8	8.628	98	211099	50.200	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.400%	
62) 4-Bromofluorobenzene	11.061	95	94031	61.393	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 122.780%	
Target Compounds						
					Qvalue	
6) Chloroethane	1.697	64	1245	1.758	ug/l	99
17) Carbon Disulfide	2.520	76	2047	0.691	ug/l #	93
64) Tetrachloroethene	9.256	164	74512	59.941	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101725\
Data File : VX048208.D
Acq On : 17 Oct 2025 13:43
Operator : JC/MD
Sample : Q3311-12DL 5X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-214DDL

Quant Time: Oct 18 00:17:04 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101725W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 18 00:06:31 2025
Response via : Initial Calibration

