

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101725\
 Data File : VX048215.D
 Acq On : 17 Oct 2025 16:13
 Operator : JC/MD
 Sample : Q3352-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-3

Quant Time: Oct 18 00:20:16 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 18 00:06:31 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	105060	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	213655	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	237995	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	124797	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	89102	54.745	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.480%	
35) Dibromofluoromethane	5.373	113	74183	50.028	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.060%	
50) Toluene-d8	8.635	98	257222	48.151	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 96.300%	
62) 4-Bromofluorobenzene	11.061	95	115487	59.355	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 118.720%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	11897	21.428	ug/l #	89
18) Methyl Acetate	2.703	43	5153	4.557	ug/l	97
20) Methylene Chloride	2.794	84	4930	3.645	ug/l	95
25) 2-Butanone	4.556	43	5650	8.735	ug/l #	89
43) Isopropyl Acetate	6.336	43	6924	2.471	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101725\
Data File : VX048215.D
Acq On : 17 Oct 2025 16:13
Operator : JC/MD
Sample : Q3352-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-3

Quant Time: Oct 18 00:20:16 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101725W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 18 00:06:31 2025
Response via : Initial Calibration

