

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101725\
 Data File : VX048219.D
 Acq On : 17 Oct 2025 17:36
 Operator : JC/MD
 Sample : Q3357-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DRUM-3

Quant Time: Oct 18 00:22:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101725W.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 18 00:06:31 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	103417	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.745	114	202917	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.037	117	217243	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.006	152	113390	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.946	65	83705	52.246	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	104.500%
35) Dibromofluoromethane	5.373	113	70043	49.736	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.480%
50) Toluene-d8	8.634	98	244052	48.103	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.200%
62) 4-Bromofluorobenzene	11.061	95	107766	58.318	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	116.640%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	14934	27.325	ug/l	97
18) Methyl Acetate	2.709	43	5060	4.546	ug/l	100
20) Methylene Chloride	2.794	84	8495	6.380	ug/l	95
25) 2-Butanone	4.568	43	6452	10.133	ug/l #	78
43) Isopropyl Acetate	6.336	43	6270	2.356	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX101725\
Data File : VX048219.D
Acq On : 17 Oct 2025 17:36
Operator : JC/MD
Sample : Q3357-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DRUM-3

Quant Time: Oct 18 00:22:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X101725W.M
Quant Title : SW846 8260
QLast Update : Sat Oct 18 00:06:31 2025
Response via : Initial Calibration

