

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101818\
 Data File : VX005433.D
 Acq On : 18 Oct 2018 16:36
 Operator : JC/MD
 Sample : J5202-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PL-03-101718-C

Quant Time: Oct 19 04:54:26 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	304838	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	446738	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	399344	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	193960	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	170246	50.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.64%	
35) Dibromofluoromethane	5.50	113	138814	46.91	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.82%	
50) Toluene-d8	8.72	98	525003	51.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.36%	
62) 4-Bromofluorobenzene	11.14	95	174984	45.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.72%	
Target Compounds						
16) Acetone	2.45	43	12847	6.568	ug/l	96
18) Methyl Acetate	2.78	43	10805	1.942	ug/l	94
25) 2-Butanone	4.71	43	3112	1.087	ug/l	99
43) Isopropyl Acetate	6.46	43	183684	24.782	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101818\
Data File : VX005433.D
Acq On : 18 Oct 2018 16:36
Operator : JC/MD
Sample : J5202-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PL-03-101718-C

Quant Time: Oct 19 04:54:26 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration

