

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101818\
 Data File : VX005436.D
 Acq On : 18 Oct 2018 17:57
 Operator : JC/MD
 Sample : J5536-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 351-SOIL

Quant Time: Oct 19 05:05:06 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	308064	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	461027	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	391070	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	189800	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	175059	51.20	ug/l	-0.01
Spiked Amount				50.000		
			Recovery	=	102.40%	
35) Dibromofluoromethane	5.50	113	142254	46.59	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	93.18%	
50) Toluene-d8	8.71	98	521934	49.79	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	99.58%	
62) 4-Bromofluorobenzene	11.14	95	173613	44.09	ug/l	0.00
Spiked Amount				50.000		
			Recovery	=	88.18%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	38185	19.316	ug/l	95
17) Carbon Disulfide	2.57	76	10669	1.305	ug/l	99
18) Methyl Acetate	2.78	43	9703	1.725	ug/l	94
25) 2-Butanone	4.71	43	6632	2.293	ug/l	89
40) Benzene	6.15	78	18954	1.415	ug/l	95
43) Isopropyl Acetate	6.46	43	168833	22.072	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101818\
Data File : VX005436.D
Acq On : 18 Oct 2018 17:57
Operator : JC/MD
Sample : J5536-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
351-SOIL

Quant Time: Oct 19 05:05:06 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration

