

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX101818\
 Data File : VX005437.D
 Acq On : 18 Oct 2018 18:23
 Operator : JC/MD
 Sample : J5590-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LS-DRUM

Quant Time: Oct 19 13:55:14 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 12 11:28:43 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	317308	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	441221	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	388323	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	189020	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	168839	47.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.88%	
35) Dibromofluoromethane	5.49	113	140374	48.04	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.08%	
50) Toluene-d8	8.71	98	518219	51.65	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.30%	
62) 4-Bromofluorobenzene	11.14	95	170234	45.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.34%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX101818\
Data File : VX005437.D
Acq On : 18 Oct 2018 18:23
Operator : JC/MD
Sample : J5590-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LS-DRUM

Quant Time: Oct 19 13:55:14 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X101218W.M
Quant Title : SW846 8260
QLast Update : Fri Oct 12 11:28:43 2018
Response via : Initial Calibration

